



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB97173

Ambient Barometric Pressure (mmHg): 755

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.0	80.6	80.9	08/08/2018	10:42
2	J4374-02		1	7	>100.0	>212.0	>212.0	08/08/2018	11:05
3	J4374-02DUP		1	8	>100.0	>212.0	>212.0	08/08/2018	11:40

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

LB 917173

WORKLIST(Hardcopy Internal Chain)

WorkList Name : fp8818

WorkList ID : 115402

Date : 8/8/2018 10:11:07 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/10/2018	Water	J4374-02	C Flash Point	Cool 4 deg C	PSEG01	K63	OILY-WATER	08/07/2018	1010A

Date/Time 08/08/18 10:15 AM
Received by: KS
Relinquished by: CP

Date/Time 08/08/18 11:45 AM
Received by: CP
Relinquished by: KS