


L397178

7471B

INSTRUMENT ID: CV1

Linear



A= 0.0000e+000

B= 2.1486e-004

C= 2.4118e-002

Rho= 0.9998511

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.042	0.042	85	0.000	85				
0.2	0.200	0.238	0.038	996	0.0 %	996				
2.5	2.500	2.483	-0.017	11442	0.0 %	11442				
5.0	5.000	4.872	-0.128	22561	0.0 %	22561				
7.5	7.500	7.500	0.000	34795	0.0 %	34795				
10.0	10.000	10.065	0.065	46734	0.0 %	46734				

LB97178

Method: 7471B

Operator: Admin

Date of Analysis: 08 Aug 2018 13:58:15

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	S0	85	-	0.0000	7471B	PPB	08 Aug 2018 14:00:08	S	Std
0.2 - 1	S0.2	996	-	0.2000	7471B	PPB	08 Aug 2018 14:02:25	S	Std
2.5 - 1	S2.5	11442	-	2.5000	7471B	PPB	08 Aug 2018 14:04:43	S	Std
5.0 - 1	S5	22561	-	5.0000	7471B	PPB	08 Aug 2018 14:07:01	S	Std
7.5 - 1	S7.5	34795	-	7.5000	7471B	PPB	08 Aug 2018 14:09:18	S	Std
10.0 - 1	S10	46734	-	10.0000	7471B	PPB	08 Aug 2018 14:11:35	S	Std
ICV96 - 1	ICV96	18632	4.0274	-	7471B	PPB	08 Aug 2018 14:14:19	U	SMPL
ICB96 - 1	ICB96	-44	0.0147	-	7471B	PPB	08 Aug 2018 14:16:35	U	SMPL
CCV32 - 1	CCV32	22663	4.8935	-	7471B	PPB	08 Aug 2018 14:18:51	U	SMPL
CCB32 - 1	CCB32	-18	0.0203	-	7471B	PPB	08 Aug 2018 14:21:07	U	SMPL
CRA - 1	CRA	937	0.2254	-	7471B	PPB	08 Aug 2018 14:23:24	U	SMPL
HighStd - 1	HighStd	46706	10.0593	-	7471B	PPB	08 Aug 2018 14:25:40	U	SMPL
ChkStd - 1	ChkStd	32051	6.9105	-	7471B	PPB	08 Aug 2018 14:27:55	U	SMPL
PB111858BL - 1	PBS	-263	-0.0324	-	7471B	PPB	08 Aug 2018 14:30:14	U	SMPL
PB111858BS - 1	LCSS	16113	3.4861	-	7471B	PPB	08 Aug 2018 14:32:33	U	SMPL
J4275-01 - 1	CL-02-080718-A	1244	0.2914	-	7471B	PPB	08 Aug 2018 14:34:50	U	SMPL
J4275-03 - 1	CL-02-080718-B	7885	1.7183	-	7471B	PPB	08 Aug 2018 14:37:07	U	SMPL
J4275-05 - 1	CL-02-080718-C	2157	0.4876	-	7471B	PPB	08 Aug 2018 14:39:25	U	SMPL
J4275-07 - 1	CL-02-080718-D	3903	0.8627	-	7471B	PPB	08 Aug 2018 14:41:42	U	SMPL
J4276-01 - 1	OK-02-080718	3929	0.8683	-	7471B	PPB	08 Aug 2018 14:43:58	U	SMPL
CCV33 - 1	CCV33	21743	4.6958	-	7471B	PPB	08 Aug 2018 14:46:14	U	SMPL
CCB33 - 1	CCB33	-63	0.0106	-	7471B	PPB	08 Aug 2018 14:48:29	U	SMPL
J4276-01DUP - 1	OK-02-080718DUP	3458	0.7671	-	7471B	PPB	08 Aug 2018 14:50:45	U	SMPL
J4276-01MS - 1	OK-02-080718MS	20667	4.4646	-	7471B	PPB	08 Aug 2018 14:53:02	U	SMPL
J4276-01MSD - 1	OK-02-080718MSD	21246	4.5890	-	7471B	PPB	08 Aug 2018 14:55:18	U	SMPL
J4375-01 - 1	OILY-DEBRIS-DRUM-A	-15	0.0209	-	7471B	PPB	08 Aug 2018 14:57:35	U	SMPL
J4375-03 - 1	OILY-DEBRIS-DRUM-B	398	0.1096	-	7471B	PPB	08 Aug 2018 14:59:52	U	SMPL
J4276-01LX5 - 1		721	0.1790	-	7471B	PPB	08 Aug 2018 15:02:09	U	SMPL
J4276-01A - 1		21668	4.6797	-	7471B	PPB	08 Aug 2018 15:04:27	U	SMPL
CCV34 - 1	CCV34	22093	4.7710	-	7471B	PPB	08 Aug 2018 15:06:44	U	SMPL
CCB34 - 1	CCB34	-40	0.0155	-	7471B	PPB	08 Aug 2018 15:09:02	U	SMPL