


245.1

LB 97184
INSTRUMENT ID: CV2

Linear

 μ Abs.:

32443



A= 0.0000e+000

B= 3.0162e-004

C= -2.5052e-002

Rho= 0.9992645

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.043	-0.043	-58	0.000	-58				
0.05	0.050	0.055	0.005	264	0.0 %	264				
0.2	0.200	0.190	-0.010	714	0.0 %	714				
2.5	2.500	2.447	-0.053	8197	0.0 %	8197				
5.0	5.000	5.065	0.065	16875	0.0 %	16875				
7.5	7.500	7.775	0.275	25860	0.0 %	25860				
10.0	10.000	9.761	-0.239	32443	0.0 %	32443				

LB97184 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 08 Aug 2018 15:41:11

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-58	-	0.0000	245.1	PPB	08 Aug 2018 15:46:26	S	Std
0.05 - 1	50.05	264	-	0.0500	245.1	PPB	08 Aug 2018 15:48:34	S	Std
0.2 - 1	50.2	714	-	0.2000	245.1	PPB	08 Aug 2018 15:50:43	S	Std
2.5 - 1	52.5	8197	-	2.5000	245.1	PPB	08 Aug 2018 15:52:51	S	Std
5.0 - 1	55	16875	-	5.0000	245.1	PPB	08 Aug 2018 15:54:58	S	Std
7.5 - 1	57.5	25860	-	7.5000	245.1	PPB	08 Aug 2018 15:57:05	S	Std
10.0 - 1	510	32443	-	10.0000	245.1	PPB	08 Aug 2018 16:01:39	S	Std
ICV97 - 1	ICV97	13524	4.0541	-245.1		PPB	08 Aug 2018 16:06:06	U	SMPL
ICB97 - 1	ICB97	-166	-0.0751	-245.1		PPB	08 Aug 2018 16:08:12	U	SMPL
CCV35 - 1	CCV35	16931	5.0817	-245.1		PPB	08 Aug 2018 16:10:19	U	SMPL
CCB35 - 1	CCB35	-177	-0.0784	-245.1		PPB	08 Aug 2018 16:12:26	U	SMPL
CRA - 1	CRA	230	0.0443	-245.1		PPB	08 Aug 2018 16:14:34	U	SMPL
HighStd - 1	HighStd	32037	9.6381	-245.1		PPB	08 Aug 2018 16:16:41	U	SMPL
ChkStd - 1	ChkStd	22714	6.8260	-245.1		PPB	08 Aug 2018 16:18:48	U	SMPL
PB111876BL - 1	PBW	-56	-0.0419	-245.1		PPB	08 Aug 2018 16:23:23	U	SMPL
PB111876BS - 1	LCSW	13515	4.0514	-245.1		PPB	08 Aug 2018 16:25:30	U	SMPL
J4350-01 - 1	080218-DBRI-E-D-B	-149	-0.0700	-245.1		PPB	08 Aug 2018 16:27:38	U	SMPL
J4350-02 - 1	080218-DBRI-E-D-S	-46	-0.0389	-245.1		PPB	08 Aug 2018 16:29:46	U	SMPL
J4350-03 - 1	080218-DBRI-E-U-B	-41	-0.0374	-245.1		PPB	08 Aug 2018 16:31:54	U	SMPL
J4350-04 - 1	080218-DBRI-E-U-S	-37	-0.0362	-245.1		PPB	08 Aug 2018 16:34:03	U	SMPL
J4350-05 - 1	080218-TIPC-E-D-M	-49	-0.0398	-245.1		PPB	08 Aug 2018 16:36:09	U	SMPL
CCV36 - 1	CCV36	16840	5.0543	-245.1		PPB	08 Aug 2018 16:38:16	U	SMPL
CCB36 - 1	CCB36	-189	-0.0821	-245.1		PPB	08 Aug 2018 16:40:22	U	SMPL
J4350-06 - 1	080218-TIPC-E-U-M	-71	-0.0465	-245.1		PPB	08 Aug 2018 16:42:31	U	SMPL
J4378-01 - 1	080618-TIPC-E-D-M	-78	-0.0486	-245.1		PPB	08 Aug 2018 16:44:38	U	SMPL
J4378-02 - 1	080618-TIPC-E-U-M	-49	-0.0398	-245.1		PPB	08 Aug 2018 16:46:45	U	SMPL
J4378-02DUP - 1	080618-TIPC-E-U-MDUP	-76	-0.0480	-245.1		PPB	08 Aug 2018 16:48:51	U	SMPL
J4378-02MS - 1	080618-TIPC-E-U-MMS	15270	4.5807	-245.1		PPB	08 Aug 2018 16:50:59	U	SMPL
J4378-02MSD - 1	080618-TIPC-E-U-MMSD	15136	4.5403	-245.1		PPB	08 Aug 2018 16:53:07	U	SMPL
J4378-02LX5 - 1		-205	-0.0869	-245.1		PPB	08 Aug 2018 16:55:15	U	SMPL
J4378-02A - 1		13521	4.0532	-245.1		PPB	08 Aug 2018 16:57:23	U	SMPL
CCV37 - 1	CCV37	16907	5.0745	-245.1		PPB	08 Aug 2018 16:59:32	U	SMPL
CCB37 - 1	CCB37	-227	-0.0935	-245.1		PPB	08 Aug 2018 17:01:38	U	SMPL