



Analytical Summary Report

Analysis Method: 9040C

Analyst By : Sweta

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB97190

Slope : 98.7

BalanceID: WC SC-4

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			22.0	4.01	08/08/2018	17:25
2	CAL2	1	Water			22.1	7.00	08/08/2018	17:27
3	CAL3	1	Water			21.9	10.04	08/08/2018	17:29
4	ICV	1	Water			22.2	7.00	08/08/2018	17:31
5	CCV1	1	Water			22.4	2.00	08/08/2018	17:33
6	J4271-02	1	Solid	20.02	20	25.3	9.02	08/08/2018	17:35
7	J4271-02DUP	1	Solid	20.03	20	25.4	9.03	08/08/2018	17:37
8	J4271-04	1	Solid	20.02	20	25.2	9.00	08/08/2018	17:39
9	J4271-06	1	Solid	20.02	20	25.6	9.04	08/08/2018	17:41
10	J4392-01	1	Solid	20.04	20	24.6	8.00	08/08/2018	17:43
11	CCV2	1	Water			22.0	12.00	08/08/2018	17:45

WORKLIST(Hardcopy Internal Chain)

LB97190

WorkList Name : corrosivity080818

WorkList ID : 115415

Date : 8/8/2018 3:49:36 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/08/2018	Solid	J4271-02	Corrosivity	Cool 4 deg C	PSEG05	K63	HR-05-080818-A	08/08/2018	9040C
08/08/2018	Solid	J4271-04	Corrosivity	Cool 4 deg C	PSEG05	K63	HR-05-080818-B	08/08/2018	9040C
08/08/2018	Solid	J4271-06	Corrosivity	Cool 4 deg C	PSEG05	K63	HR-05-080818-C	08/08/2018	9040C
08/08/2018	Solid	J4392-01	Corrosivity	Cool 4 deg C	PARS15	P13	JTY-WC-S-1-1-62	08/08/2018	9040C

5:15 SP

Date/Time 08/09/18 4:00
 Received by: SP
 Relinquished by: CP

Date/Time 08/08/18 5:55
 Received by: CP
 Relinquished by: SP