


7471B

LB 97197
INSTRUMENT ID: CV1

Linear

 μ Abs.:

44872



A= 0.0000e+000

B= 2.2086e-004

C= -1.8774e-004

Rho= 0.9998153

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.010	0.010	46	0.000	46				
0.2	0.200	0.203	0.003	918	0.0 %	918				
2.5	2.500	2.446	-0.054	11076	0.0 %	11076				
5.0	5.000	4.993	-0.007	22607	0.0 %	22607				
7.5	7.500	7.638	0.138	34584	0.0 %	34584				
10.0	10.000	9.910	-0.090	44872	0.0 %	44872				

LB97197 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 09 Aug 2018 14:12:41

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	46	-	0.0000	7471B	PPB	09 Aug 2018 14:15:07	S	Std
0.2 - 1	50.2	918	-	0.2000	7471B	PPB	09 Aug 2018 14:17:25	S	Std
2.5 - 1	52.5	11076	-	2.5000	7471B	PPB	09 Aug 2018 14:19:42	S	Std
5.0 - 1	55	22607	-	5.0000	7471B	PPB	09 Aug 2018 14:21:59	S	Std
7.5 - 1	57.5	34584	-	7.5000	7471B	PPB	09 Aug 2018 14:24:16	S	Std
10.0 - 1	510	44872	-	10.0000	7471B	PPB	09 Aug 2018 14:29:11	S	Std
ICV99 - 1	ICV99	18108	3.9992	-	7471B	PPB	09 Aug 2018 14:32:00	U	SMPL
ICB99 - 1	ICB99	-138	-0.0307	-	7471B	PPB	09 Aug 2018 14:34:15	U	SMPL
CCV43 - 1	CCV43	22545	4.9792	-	7471B	PPB	09 Aug 2018 14:36:31	U	SMPL
CCB43 - 1	CCB43	-139	-0.0309	-	7471B	PPB	09 Aug 2018 14:38:46	U	SMPL
CRA - 1	CRA	934	0.2061	-	7471B	PPB	09 Aug 2018 14:41:04	U	SMPL
HighStd - 1	HighStd	45709	10.0952	-	7471B	PPB	09 Aug 2018 14:43:20	U	SMPL
ChkStd - 1	ChkStd	32679	7.2174	-	7471B	PPB	09 Aug 2018 14:45:36	U	SMPL
PB111890BL - 1	PBS	-224	-0.0497	-	7471B	PPB	09 Aug 2018 14:47:55	U	SMPL
PB111890BS - 1	LCSS	16405	3.6231	-	7471B	PPB	09 Aug 2018 14:50:13	U	SMPL
J4271-01 - 1	HR-05-080818-A	11341	2.5046	-	7471B	PPB	09 Aug 2018 14:52:30	U	SMPL
J4271-03 - 1	HR-05-080818-B	12179	2.6897	-	7471B	PPB	09 Aug 2018 14:54:47	U	SMPL
J4271-05 - 1	HR-05-080818-C	11238	2.4819	-	7471B	PPB	09 Aug 2018 14:57:05	U	SMPL
J4382-04 - 1	JTY-SB-03-04-05-1-11	1259	0.2779	-	7471B	PPB	09 Aug 2018 14:59:23	U	SMPL
J4382-04DUP - 1	JTY-SB-03-04-05-1-11DUP	1272	0.2807	-	7471B	PPB	09 Aug 2018 15:01:39	U	SMPL
CCV44 - 1	CCV44	21975	4.8533	-	7471B	PPB	09 Aug 2018 15:03:55	U	SMPL
CCB44 - 1	CCB44	-144	-0.0320	-	7471B	PPB	09 Aug 2018 15:06:11	U	SMPL
J4382-05 - 1	J4382-04MS	18554	4.0977	-	7471B	PPB	09 Aug 2018 15:08:26	U	SMPL
J4382-06 - 1	J4382-04MSD	17990	3.9731	-	7471B	PPB	09 Aug 2018 15:10:43	U	SMPL
J4382-07 - 1	JTY-SB-103-04-05-1-11	1759	0.3883	-	7471B	PPB	09 Aug 2018 15:12:59	U	SMPL
J4388-01 - 1	S-1A	741	0.1635	-	7471B	PPB	09 Aug 2018 15:15:15	U	SMPL
J4388-02 - 1	S-1B	560	0.1235	-	7471B	PPB	09 Aug 2018 15:17:33	U	SMPL
J4388-03 - 1	S-2A	1361	0.3004	-	7471B	PPB	09 Aug 2018 15:19:50	U	SMPL
J4388-04 - 1	S-2B	2114	0.4667	-	7471B	PPB	09 Aug 2018 15:22:06	U	SMPL
J4388-05 - 1	S-3A	1920	0.4239	-	7471B	PPB	09 Aug 2018 15:24:24	U	SMPL
J4388-06 - 1	S-3B	588	0.1297	-	7471B	PPB	09 Aug 2018 15:26:41	U	SMPL
J4388-07 - 1	CHA-1	2481	0.5478	-	7471B	PPB	09 Aug 2018 15:28:57	U	SMPL
CCV45 - 1	CCV45	22422	4.9520	-	7471B	PPB	09 Aug 2018 15:31:13	U	SMPL
CCB45 - 1	CCB45	-120	-0.0267	-	7471B	PPB	09 Aug 2018 15:33:29	U	SMPL
J4393-03 - 1	JTY-SB-01-1-62	711	0.1568	-	7471B	PPB	09 Aug 2018 15:35:46	U	SMPL
J4393-04 - 1	JTY-BALLARDS-1-2	2453	0.5416	-	7471B	PPB	09 Aug 2018 15:38:02	U	SMPL
J4394-01 - 1	FK-EF-01-006-ABCD	3877	0.8561	-	7471B	PPB	09 Aug 2018 15:40:18	U	SMPL
J4395-01 - 1	FK-BPM-01-E7-F7-E7A-F7A	1387	0.3061	-	7471B	PPB	09 Aug 2018 15:42:35	U	SMPL
J4396-01 - 1	FK-BPM-01-E8-F8-E8A-F8A	1656	0.3656	-	7471B	PPB	09 Aug 2018 15:44:52	U	SMPL
J4398-01 - 1	EPE-P10A(0-5)	12117	2.6760	-	7471B	PPB	09 Aug 2018 15:47:09	U	SMPL
J4398-02 - 1	EPE-P10A(5-10)	4754	1.0498	-	7471B	PPB	09 Aug 2018 15:49:26	U	SMPL
J4398-03 - 1	EPE-P10A(37-42)	685	0.1511	-	7471B	PPB	09 Aug 2018 15:51:43	U	SMPL
J4382-04LX5 - 1		242	0.0533	-	7471B	PPB	09 Aug 2018 15:54:01	U	SMPL
J4382-04A - 1		18927	4.1801	-	7471B	PPB	09 Aug 2018 15:56:17	U	SMPL
CCV46 - 1	CCV46	22665	5.0057	-	7471B	PPB	09 Aug 2018 15:58:34	U	SMPL
CCB46 - 1	CCB46	-117	-0.0260	-	7471B	PPB	09 Aug 2018 16:00:50	U	SMPL