



Analytical Summary Report

Analysis Method: 7196A
Parameter: Hexavalent Chromium
Run Number: LB97199

ANALYST: Indiana
SUPERVISOR REVIEW BY: apatel
pH Meter ID: WC pH Meter-1

Table with 2 columns: Reagent/Standard, Lot/Log #. Rows include Calibration Std. hexchrome 0.1 ppm, 5N sulfuric acid, Hexavalent Chromium ICV Std., etc.

Intercept: 0.0002 Slope: 0.7677 Regression: 0.999975

Table with 12 columns: Seq, Lab ID, True Value (mg/l), DF, Initial Vol (ml/gm), Final Vol (ml), pH HN03, pH H2SO4, Absorb. at 540 (Backgrnd, Color), Absorbance Difference, Anal Date, Anal Time. Contains 7 rows of calibration data.



Analytical Summary Report

Analysis Method: 7196A

ANALYST: Indiana

Parameter: Hexavalent Chromium

SUPERVISOR REVIEW BY: apatel

Run Number: LB97199

pH Meter ID: WC pH Meter-1

Seq	Lab ID	True Value (mg/l)	DF	Initial Vol (ml/gm)	Final Vol (ml)	pH		Absorb. at 540		Absorbance Difference	Result	Anal Date	Anal Time
						HN03	H2SO4	Backgrnd	Color				
1	ICV	0.5	1	100	100		1.55	0.000	0.392	0.392	0.510	08/09/2018	11:43
2	ICB		1	100	100		2.13	0.000	0.000	0.000	0.000	08/09/2018	11:44
3	CCV1	0.5	1	100	100		2.42	0.000	0.393	0.393	0.512	08/09/2018	11:44
4	CCB1		1	100	100		1.88	0.000	0.000	0.000	0.000	08/09/2018	11:45
5	PB111907BL		1	2.50	100	7.64	1.74	0.000	0.000	0.000	0.000	08/09/2018	11:45
6	PB111907BS	20	1	2.54	100	7.36	2.02	0.000	0.381	0.381	0.496	08/09/2018	11:46
7	J4271-01		1	2.52	100	7.22	2.40	0.014	0.014	0.000	0.000	08/09/2018	11:46
8	J4271-03		1	2.55	100	7.11	2.29	0.018	0.018	0.000	0.000	08/09/2018	11:47
9	J4271-05		1	2.50	100	7.38	1.63	0.020	0.020	0.000	0.000	08/09/2018	11:47
10	J4382-04		1	2.55	100	7.21	1.92	0.007	0.031	0.024	0.031	08/09/2018	11:48
11	J4382-04DUP		1	2.53	100	7.25	2.03	0.008	0.032	0.024	0.031	08/09/2018	11:48
12	J4382-05Pre	40	2	2.51	100	7.14	1.77	0.010	0.369	0.359	0.467	08/09/2018	11:49
13	J4382-052Ins	1284	40	2.51	100	7.10	1.80	0.006	0.612	0.606	0.789	08/09/2018	11:49
14	J4382-053Post	40	2	2.51	100	7.19	2.24	0.011	0.377	0.366	0.476	08/09/2018	11:50
15	CCV2	0.5	1	100	100		2.32	0.000	0.393	0.393	0.512	08/09/2018	11:50
16	CCB2		1	100	100		1.57	0.000	0.000	0.000	0.000	08/09/2018	11:51
17	J4382-06	40	2	2.51	100	7.25	2.03	0.010	0.369	0.359	0.467	08/09/2018	11:52
18	J4382-07		1	2.54	100	7.65	1.91	0.009	0.028	0.019	0.024	08/09/2018	11:52
19	J4393-03		1	2.50	100	7.23	2.15	0.038	0.038	0.000	0.000	08/09/2018	11:52
20	J4393-04		1	2.51	100	7.80	2.43	0.015	0.254	0.239	0.311	08/09/2018	11:53
21	J4394-01		1	2.52	100	7.96	1.98	0.017	0.017	0.000	0.000	08/09/2018	11:53
22	J4395-01		1	2.53	100	7.88	1.81	0.013	0.013	0.000	0.000	08/09/2018	11:54
23	J4396-01		1	2.55	100	7.71	1.95	0.010	0.010	0.000	0.000	08/09/2018	11:54
24	CCV3	0.5	1	100	100		2.10	0.000	0.393	0.393	0.512	08/09/2018	11:55
25	CCB3		1	100	100		1.66	0.000	0.000	0.000	0.000	08/09/2018	11:55