



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: AMANDEEP

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB97200

Reagent/Standard	Lot/Loq #
TURBIDITY CCV 10 NTU	WP66871
Turbidity std, 80NTU	WP66864
Turbidity std, 40NTU	WP66865
turbidity std 5 NTU	WP66868
turbidity std 20 NTU	WP66866
Turbidity std 1 NTU	WP66869
turbidity std 0 NTU	WP66870
turbidity std 10 NTU	WP66867
10 NTU Standard 500 ml	W2297

Intercept: 0.3762

Slope: 0.8692

Regression: 0.9998

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Anal Date	Anal Time
1	CAL1	0	1	0.229	08/09/2018	14:05
2	CAL2	1	1	0.996	08/09/2018	14:10
3	CAL3	5	1	4.860	08/09/2018	14:15
4	CAL4	10	1	9.880	08/09/2018	14:20
5	CAL5	20	1	16.860	08/09/2018	14:24
6	CAL6	40	1	35.600	08/09/2018	14:29
7	CAL7	80	1	69.800	08/09/2018	14:31



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: AMANDEEP

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB97200

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (NTU)	AnalDate	Anal Time
1	ICV	10	1	9.260	10.221	08/09/2018	14:34
2	ICB		1	0.254	-0.141	08/09/2018	14:38
3	CCV1	10	1	9.460	10.451	08/09/2018	14:43
4	CCB1		1	0.216	-0.184	08/09/2018	14:46
5	LB97200BL		1	0.220	-0.180	08/09/2018	14:52
6	J4407-01		1	45.200	51.569	08/09/2018	14:56
7	J4407-01DUP		1	47.100	53.755	08/09/2018	14:59
8	CCV2	10	1	9.360	10.336	08/09/2018	15:03
9	CCB2		1	0.219	-0.181	08/09/2018	15:09

LB97200

WORKLIST(Hardcopy Internal Chain)

WorkList Name : tur080918

WorkList ID : 115455

Date : 8/9/2018 12:00:13 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/10/2018	Water	J4407-01	Turbidity	Cool 4 deg C	LIRO01	O51	UDS-CH-4-01	08/08/2018	SM2130 B

Date/Time 8-9-18 13:45
Received by: AR
Relinquished by: JP

Date/Time 8-9-18 15:10
Received by: JP
Relinquished by: AD