

ISM02.4


LB97213

INSTRUMENT ID: CV2

Linear

 μ Abs.:

41354



A= 0.0000e+000

B= 2.4103e-004

C= 1.5249e-002

Rho= 0.9998676

Accept=Accepted

slope
y-intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.018	0.018	13	0.000	13					-
0.2	0.200	0.224	0.024	865	0.0 %	865					12
2.5	2.500	2.486	-0.014	10253	0.0 %	10253					-1
5.0	5.000	4.894	-0.106	20241	0.0 %	20241					-2
7.5	7.500	7.595	0.095	31447	0.0 %	31447					1
10.0	10.000	9.983	-0.017	41354	0.0 %	41354					0

103
8/10/18

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0.0	13	-	0	ISM02.4	PPB	8/10/2018 9:34	Std
	0.2 S0.2	865	-	0.2	ISM02.4	PPB	8/10/2018 9:36	Std
	2.5 S2.5	10253	-	2.5	ISM02.4	PPB	8/10/2018 9:39	Std
	5 S5	20241	-	5	ISM02.4	PPB	8/10/2018 9:41	Std
	7.5 S7.5	31447	-	7.5	ISM02.4	PPB	8/10/2018 9:43	Std
	10 S10.0	41354	-	10	ISM02.4	PPB	8/10/2018 9:51	Std
ICV	ICV	16668	4.0327 -		ISM02.4	PPB	8/10/2018 9:54	SMPL
ICB	ICB	-185	-0.0293 -		ISM02.4	PPB	8/10/2018 9:56	SMPL
CCV047	CCV047	20950	5.0648 -		ISM02.4	PPB	8/10/2018 9:58	SMPL
CCB047	CCB047	-149	-0.0207 -		ISM02.4	PPB	8/10/2018 10:00	SMPL
PB111918BL	PBW003	-44	0.0046 -		ISM02.4	PPB	8/10/2018 10:03	SMPL
J4404-01	MC0AD6	-14	0.0119 -		ISM02.4	PPB	8/10/2018 10:05	SMPL
J4404-02	MC0AD7	-67	-0.0009 -		ISM02.4	PPB	8/10/2018 10:07	SMPL
J4404-03	MC0AD8	-93	-0.0072 -		ISM02.4	PPB	8/10/2018 10:09	SMPL
J4404-04	MC0AD9	-59	0.001 -		ISM02.4	PPB	8/10/2018 10:12	SMPL
J4404-05	MC0AD9D	-57	0.0015 -		ISM02.4	PPB	8/10/2018 10:14	SMPL
J4404-06	MC0AD9S	4183	1.0235 -		ISM02.4	PPB	8/10/2018 10:16	SMPL
J4404-07	MC0AE0	-114	-0.0122 -		ISM02.4	PPB	8/10/2018 10:19	SMPL
J4404-08	MC0AE1	-59	0.001 -		ISM02.4	PPB	8/10/2018 10:21	SMPL
CCV048	CCV048	20778	5.0233 -		ISM02.4	PPB	8/10/2018 10:23	SMPL
CCB048	CCB048	-137	-0.0178 -		ISM02.4	PPB	8/10/2018 10:25	SMPL