

Analytical Summary ReportAnalysis Method: 365.3ANALYST: AparnaParameter: Phosphorus-TotalSUPERVISOR REVIEW BY: apatelRun Number: LB97217

Reagent/Standard	Lot/Log #
5N sulfuric acid	WP65054
calibration std. phosphate 1 ppm	WP66900
calibration std. phosphate 0.5 ppm	WP66899
calibration std. phosphate 0.3 ppm	WP66898
calibration std. 0 ppm	WP66895
phosphate ICV std.	WP66902
phosphate CCV std.	WP66901
calibration std. phosphate 0.1 ppm	WP66897
calibration std. phosphate 0.05 ppm	WP66896
Combined reagent	WP66904
11 N sulfuric acid	WP66905
Phenolphthalein indicator	WP66907
phosphate ICV std.	WP66902
phosphate CCV std.	WP66901

Intercept: -0.001Slope: 0.6502Regression: 0.999847

Seq	Lab ID	True Value (mg/L)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	AnalDate	AnalTime
1	CAL1	0.00	1	50	50	0.000	08/10/2018	13:03
2	CAL2	0.05	1	50	50	0.033	08/10/2018	13:06
3	CAL3	0.10	1	50	50	0.065	08/10/2018	13:06
4	CAL4	0.30	1	50	50	0.186	08/10/2018	13:07
5	CAL5	0.50	1	50	50	0.329	08/10/2018	13:08
6	CAL6	1.00	1	50	50	0.649	08/10/2018	13:09

Analysis Method: 365.3ANALYST: AparnaParameter: Phosphorus-TotalSUPERVISOR REVIEW BY: apatelRun Number: LB97217

Seq	Lab ID	True Value (mg/l)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	Result (mg/L)	AnalDate	AnalTime
1	ICV	0.50	1	50	50	0.325	0.501	08/10/2018	13:09
2	ICB		1	50	50	0.000	0.002	08/10/2018	13:10
3	CCV1	0.50	1	50	50	0.327	0.504	08/10/2018	13:10
4	CCB1		1	50	50	0.000	0.002	08/10/2018	13:11
5	PB111958BL		1	50	50	0.000	0.002	08/10/2018	13:12
6	PB111958BS	0.50	1	50	50	0.302	0.466	08/10/2018	13:12
7	J4403-01		1	50	50	0.661	1.018	08/10/2018	13:13
8	J4403-01DUP		1	50	50	0.661	1.018	08/10/2018	13:13
9	J4403-01MS	0.50	1	50	50	0.855	1.317	08/10/2018	13:14
10	J4403-01MSD	0.50	1	50	50	0.879	1.353	08/10/2018	13:14
11	J4403-01		5	50	50	0.135	0.209	08/10/2018	13:20
12	J4403-01DUP		5	50	50	0.141	0.218	08/10/2018	13:20
13	J4403-01MS	0.50	5	50	50	0.215	0.332	08/10/2018	13:21
14	J4403-01MSD	0.50	5	50	50	0.210	0.325	08/10/2018	13:22
15	CCV2	0.50	1	50	50	0.326	0.503	08/10/2018	13:23
16	CCB2		1	50	50	0.000	0.002	08/10/2018	13:26