

LB 97219


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

35350



A= 0.0000e+000

B= 2.8078e-004

C= -2.4653e-002

Rho= 0.9997393

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.029	0.029	190	0.000	190				
0.2	0.200	0.176	-0.024	713	0.0 %	713				
2.5	2.500	2.477	-0.023	8909	0.0 %	8909				
5.0	5.000	4.950	-0.050	17716	0.0 %	17716				
7.5	7.500	7.668	0.168	27398	0.0 %	27398				
10.0	10.000	9.901	-0.099	35350	0.0 %	35350				

LB97219 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 10 Aug 2018 12:09:09

Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	S0	190	-	0.00007470A	PPB		10 Aug 2018 12:11:41	S	Std
0.2 - 1	S0.2	713	-	0.20007470A	PPB		10 Aug 2018 12:13:50	S	Std
2.5 - 1	S2.5	8909	-	2.50007470A	PPB		10 Aug 2018 12:15:58	S	Std
5.0 - 1	S5	17716	-	5.00007470A	PPB		10 Aug 2018 12:18:07	S	Std
7.5 - 1	S7.5	27398	-	7.50007470A	PPB		10 Aug 2018 12:20:15	S	Std
10.0 - 1	S10	35350	-	10.00007470A	PPB		10 Aug 2018 12:22:23	S	Std
ICV101 - 1	ICV101	14053	3.9212	-7470A	PPB		10 Aug 2018 12:25:00	U	SMPL
ICB101 - 1	ICB101	15	-0.0204	-7470A	PPB		10 Aug 2018 12:27:07	U	SMPL
CCV49 - 1	CCV49	18013	5.0331	-7470A	PPB		10 Aug 2018 12:29:14	U	SMPL
CCB49 - 1	CCB49	-6	-0.0263	-7470A	PPB		10 Aug 2018 12:31:20	U	SMPL
CRA - 1	CRA	709	0.1744	-7470A	PPB		10 Aug 2018 12:33:28	U	SMPL
HighStd - 1	HighStd	35701	9.9996	-7470A	PPB		10 Aug 2018 12:35:34	U	SMPL
ChkStd - 1	ChkStd	25005	6.9963	-7470A	PPB		10 Aug 2018 12:37:41	U	SMPL
PB111914BL - 1	PBW	70	-0.0050	-7470A	PPB		10 Aug 2018 12:43:00	U	SMPL
PB111914BS - 1	LCSW	13990	3.9035	-7470A	PPB		10 Aug 2018 12:45:07	U	SMPL
PB111914TB - 1	PB111887TB	-175	-0.0738	-7470A	PPB		10 Aug 2018 12:47:14	U	SMPL
J4271-02 - 1	HR-05-080818-A	-61	-0.0418	-7470A	PPB		10 Aug 2018 12:49:23	U	SMPL
J4271-02DUP - 1	HR-05-080818-ADUP	-36	-0.0348	-7470A	PPB		10 Aug 2018 12:51:31	U	SMPL
J4271-02MS - 1	HR-05-080818-AMS	13676	3.8153	-7470A	PPB		10 Aug 2018 12:53:40	U	SMPL
J4271-02MSD - 1	HR-05-080818-AMSD	13600	3.7940	-7470A	PPB		10 Aug 2018 12:55:47	U	SMPL
CCV50 - 1	CCV50	17971	5.0213	-7470A	PPB		10 Aug 2018 12:57:54	U	SMPL
CCB50 - 1	CCB50	2	-0.0241	-7470A	PPB		10 Aug 2018 13:00:02	U	SMPL
J4271-04 - 1	HR-05-080818-B	-38	-0.0353	-7470A	PPB		10 Aug 2018 13:02:10	U	SMPL
J4271-06 - 1	HR-05-080818-C	9	-0.0221	-7470A	PPB		10 Aug 2018 13:04:17	U	SMPL
J4388-08 - 1	COMP-1	0	-0.0247	-7470A	PPB		10 Aug 2018 13:06:24	U	SMPL
J4392-01 - 1	JTY-WC-S-1-1-62	-14	-0.0286	-7470A	PPB		10 Aug 2018 13:08:30	U	SMPL
PB111915BL - 1	PBW	-12	-0.0280	-7470A	PPB		10 Aug 2018 13:10:38	U	SMPL
PB111915BS - 1	LCSW	14089	3.9313	-7470A	PPB		10 Aug 2018 13:12:45	U	SMPL
J4379-07 - 1	FB-08082018	-177	-0.0744	-7470A	PPB		10 Aug 2018 13:14:53	U	SMPL
J4379-08 - 1	MW-106D	189	0.0284	-7470A	PPB		10 Aug 2018 13:17:03	U	SMPL
J4379-09 - 1	MW-105S	51	-0.0103	-7470A	PPB		10 Aug 2018 13:19:11	U	SMPL
J4379-10 - 1	MW-105I	47	-0.0115	-7470A	PPB		10 Aug 2018 13:21:18	U	SMPL
CCV51 - 1	CCV51	17948	5.0148	-7470A	PPB		10 Aug 2018 13:23:25	U	SMPL
CCB51 - 1	CCB51	2	-0.0241	-7470A	PPB		10 Aug 2018 13:25:32	U	SMPL
J4379-11 - 1	MW-105D	-86	-0.0488	-7470A	PPB		10 Aug 2018 13:27:40	U	SMPL
J4379-12 - 1	MW-103S	140	0.0147	-7470A	PPB		10 Aug 2018 13:29:47	U	SMPL
J4379-12DUP - 1	MW-103SDUP	100	0.0034	-7470A	PPB		10 Aug 2018 13:31:53	U	SMPL
J4379-13 - 1	J4379-12MS	13965	3.8965	-7470A	PPB		10 Aug 2018 13:34:00	U	SMPL
J4379-14 - 1	J4379-12MSD	14234	3.9720	-7470A	PPB		10 Aug 2018 13:36:08	U	SMPL
J4379-15 - 1	MW-104I	-169	-0.0721	-7470A	PPB		10 Aug 2018 13:38:16	U	SMPL
J4379-16 - 1	MW-104D	429	0.0958	-7470A	PPB		10 Aug 2018 13:40:25	U	SMPL
J4379-17 - 1	MW-103I	116	0.0079	-7470A	PPB		10 Aug 2018 13:42:33	U	SMPL
CCV52 - 1	CCV52	17565	4.9073	-7470A	PPB		10 Aug 2018 13:44:42	U	SMPL
CCB52 - 1	CCB52	-46	-0.0376	-7470A	PPB		10 Aug 2018 13:46:49	U	SMPL
J4271-02LX5 - 1		69	-0.0053	-7470A	PPB		10 Aug 2018 13:48:58	U	SMPL
J4271-02A - 1		13953	3.8931	-7470A	PPB		10 Aug 2018 13:51:05	U	SMPL
J4379-12LX5 - 1		-25	-0.0317	-7470A	PPB		10 Aug 2018 13:53:12	U	SMPL
J4379-12A - 1		15911	4.4429	-7470A	PPB		10 Aug 2018 13:55:20	U	SMPL
CCV53 - 1	CCV53	17738	4.9559	-7470A	PPB		10 Aug 2018 13:57:27	U	SMPL
CCB53 - 1	CCB53	-29	-0.0328	-7470A	PPB		10 Aug 2018 13:59:36	U	SMPL