



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: AMANDEEP

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB97224

Reagent/Standard	Lot/Log #
TURBIDITY CCV 10 NTU	WP66930
Turbidity std, 40NTU	WP66925
turbidity std 5 NTU	WP66928
turbidity std 20 NTU	WP66926
Turbidity std 1 NTU	WP66929
turbidity std 0 NTU	WP66931
turbidity std 10 NTU	WP66927
10 NTU Standard 500 ml	W2297

Intercept: 0.3012

Slope: 0.877

Regression: 0.9989

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Anal Date	Anal Time
1	CAL1	0	1	0.278	08/10/2018	13:30
2	CAL2	1	1	0.991	08/10/2018	13:34
3	CAL3	5	1	4.860	08/10/2018	13:38
4	CAL4	10	1	9.850	08/10/2018	13:45
5	CAL5	20	1	16.780	08/10/2018	13:50
6	CAL6	40	1	35.700	08/10/2018	13:54



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: AMANDEEP

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB97224

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (NTU)	AnalDate	Anal Time
1	ICV	10	1	8.480	9.326	08/10/2018	13:58
2	ICB		1	0.256	-0.052	08/10/2018	14:02
3	CCV1	10	1	9.560	10.557	08/10/2018	14:07
4	CCB1		1	0.246	-0.063	08/10/2018	14:11
5	LB97224BL		1	0.256	-0.052	08/10/2018	14:15
6	J4433-01		1	7.230	7.901	08/10/2018	14:18
7	J4433-02		1	6.820	7.433	08/10/2018	14:21
8	J4433-02DUP		1	6.800	7.410	08/10/2018	14:24
9	CCV2	10	1	9.260	10.215	08/10/2018	14:27
10	CCB2		1	0.240	-0.070	08/10/2018	14:30

1304224

WORKLIST(Hardcopy Internal Chain)

WorkList Name : turbidity081018

WorkList ID : 115503

Date : 8/10/2018 11:34:17 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/11/2018	Water	J4433-01	Turbidity	Cool 4 deg C	LIRO01	P31	180809-A	08/09/2018	SM2130 B
08/11/2018	Water	J4433-02	Turbidity	Cool 4 deg C	LIRO01	P31	180809-B	08/09/2018	SM2130 B

Date/Time 08-10-18 12:45
Received by: AK
Relinquished by: GP

Date/Time 08-10-18 14:35
Received by: AK
Relinquished by: AK