

TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: apatel

ANALYST: Aparna

Date: 08/13/2018

TEMP1 IN: 103 08/13/2018 07:10 TEMP OUT: 104 08/13/2018 08:15

TEMP2 IN: 104 08/13/2018 09:00 TEMP OUT: 104 08/13/2018 10:05

TEMP3 IN: 105 08/13/2018 11:30 TEMP OUT: 105 08/13/2018 13:05

TEMP4 IN: 105 08/13/2018 13:45 TEMP OUT: 105 08/13/2018 15:20

Run Number: LB97241

BalanceID: WC SC-3

OvenID: WC Oven-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L
1	LB97241BL	LB97241BL	1.3997	1.3997	100	1.3997	1.3997	1.3997	0.0000	0
2	LB97241BS	LB97241BS	1.3994	1.3994	100	1.4551	1.4551	1.4551	0.0557	557
3	J4432-01	32220008-3	1.3958	1.3958	50	1.3980	1.3980	1.3980	0.0022	44
4	J4433-01	180809-A	1.3994	1.3994	600	1.4064	1.4064	1.4064	0.0070	11.7
5	J4433-02	180809-B	1.3983	1.3983	400	1.4041	1.4041	1.4041	0.0058	14.5
6	J4440-01	080918-TIPC-F-D-B	1.4027	1.4027	250	1.4150	1.4150	1.4150	0.0123	49.2
7	J4440-01DUP	080918-TIPC-F-D-BDUP	1.4002	1.4002	250	1.4130	1.4130	1.4130	0.0128	51.2
8	J4440-02	080918-TIPC-F-D-S	1.4008	1.4008	300	1.4126	1.4126	1.4126	0.0118	39.3
9	J4440-03	080918-TIPC-F-U-B	1.3990	1.3990	300	1.4128	1.4128	1.4128	0.0138	46
10	J4440-04	080918-TIPC-F-U-S	1.4044	1.4044	400	1.4164	1.4164	1.4164	0.0120	30
11	J4441-01	COMP	1.4104	1.4104	50	1.4169	1.4169	1.4169	0.0065	130
12	J4445-01	DSN-001A-081018	1.3993	1.3993	900	1.4088	1.4088	1.4088	0.0095	10.6



TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: apatel

ANALYST: Aparna

Date: 08/13/2018

TEMP1 IN: 103 08/13/2018 07:10 TEMP OUT: 104 08/13/2018 08:15

TEMP2 IN: 104 08/13/2018 09:00 TEMP OUT: 104 08/13/2018 10:05

TEMP3 IN: 105 08/13/2018 11:30 TEMP OUT: 105 08/13/2018 13:05

TEMP4 IN: 105 08/13/2018 13:45 TEMP OUT: 105 08/13/2018 15:20

Run Number: LB97241

BalanceID: WC SC-3

OvenID: WC Oven-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L

A = Sample Volume (ml)

B = Empty Dish Weight (g)

C = Final Empty Dish + Sample weight after 1.5 hr drying @105°C(g)

D = Weight (g)

Weight (g) = C - B

Result mg/L = $\frac{D}{A} \times 1000 \times 1000$

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TSS081318

WorkList ID : 115524

Date : 8/13/2018 7:48:22 AM

2097241

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/16/2018	Water	J4432-01 <i>ND</i>	TSS	Cool 4 deg C	PHAR02	P31	32220008-3	08/09/2018	SM2540 D
08/14/2018	Water	J4433-01 <i>ND</i>	TSS	Cool 4 deg C	LIRO01	P31	180809-A	08/09/2018	SM2540 D
08/14/2018	Water	J4433-02 <i>ND</i>	TSS	Cool 4 deg C	LIRO01	P31	180809-B	08/09/2018	SM2540 D
08/14/2018	Water	J4440-01 <i>C</i>	TSS	Cool 4 deg C	HDR103	K63	080918-TIPC-F-D-B	08/09/2018	SM2540 D
08/14/2018	Water	J4440-02 <i>C</i>	TSS	Cool 4 deg C	HDR103	K63	080918-TIPC-F-D-S	08/09/2018	SM2540 D
08/14/2018	Water	J4440-03 <i>C</i>	TSS	Cool 4 deg C	HDR103	K63	080918-TIPC-F-U-B	08/09/2018	SM2540 D
08/14/2018	Water	J4440-04 <i>C</i>	TSS	Cool 4 deg C	HDR103	K63	080918-TIPC-F-U-S	08/09/2018	SM2540 D
08/17/2018	Water	J4441-01 <i>ND</i>	TSS	Cool 4 deg C	ARAM01	K63	COMP	08/10/2018	SM2540 D
08/17/2018	Water	J4445-01 <i>D</i>	TSS	Cool 4 deg C	PKFM01	K63	DSN-001A-081018	08/10/2018	SM2540 D

FP-9823888

Date/Time 08/13/18, 08:00
 Received by: *AS*
 Relinquished by: *CS*

Date/Time 08/13/18, 12:00
 Received by: *AS*
 Relinquished by: *AS*