


7470A

LB97243

INSTRUMENT ID: CV1

 μ Abs.:

34315



Linear

A= 0.0000e+000

B= 2.8903e-004

C= 1.0695e-002

Rho= 0.9998464

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.009	0.009	-7	0.000	-7				
0.2	0.200	0.209	0.009	685	0.0 %	685				
2.5	2.500	2.464	-0.036	8488	0.0 %	8488				
5.0	5.000	4.960	-0.040	17122	0.0 %	17122				
7.5	7.500	7.630	0.130	26362	0.0 %	26362				
10.0	10.000	9.929	-0.071	34315	0.0 %	34315				

LB97243 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 13 Aug 2018 09:35:56

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-7	-	0.0007470A	-7470A	PPB	13 Aug 2018 09:41:50	S	Std
0.2 - 1	50.2	685	-	0.2007470A	-7470A	PPB	13 Aug 2018 09:43:59	S	Std
2.5 - 1	52.5	8488	-	2.5007470A	-7470A	PPB	13 Aug 2018 09:46:08	S	Std
5.0 - 1	55	17122	-	5.0007470A	-7470A	PPB	13 Aug 2018 09:48:16	S	Std
7.5 - 1	57.5	26362	-	7.5007470A	-7470A	PPB	13 Aug 2018 09:50:24	S	Std
10.0 - 1	57.5	34315	-	10.0007470A	-7470A	PPB	13 Aug 2018 09:52:31	S	Std
ICV03 - 1	ICV03	13766	3.9895	-7470A	-7470A	PPB	13 Aug 2018 09:55:05	U	SMPL
ICB03 - 1	ICB03	-8	0.0084	-7470A	-7470A	PPB	13 Aug 2018 09:57:11	U	SMPL
CCV59 - 1	CCV59	17904	5.1856	-7470A	-7470A	PPB	13 Aug 2018 09:59:18	U	SMPL
CCB59 - 1	CCB59	14	0.0147	-7470A	-7470A	PPB	13 Aug 2018 10:01:25	U	SMPL
CRA - 1	CRA	644	0.1968	-7470A	-7470A	PPB	13 Aug 2018 10:03:32	U	SMPL
HighStd - 1	HighStd	34162	9.8847	-7470A	-7470A	PPB	13 Aug 2018 10:05:39	U	SMPL
ChkStd - 1	ChkStd	24271	7.0258	-7470A	-7470A	PPB	13 Aug 2018 10:07:46	U	SMPL
PB111954BL - 1	PBW	-142	-0.0303	-7470A	-7470A	PPB	13 Aug 2018 10:09:55	U	SMPL
PB111954BS - 1	LCSW	13137	3.8077	-7470A	-7470A	PPB	13 Aug 2018 10:12:04	U	SMPL
PB111954TB - 1	PB111920TB	-25	0.0035	-7470A	-7470A	PPB	13 Aug 2018 10:14:11	U	SMPL
J4272-02 - 1	JC-03-080918-A	116	0.0442	-7470A	-7470A	PPB	13 Aug 2018 10:16:20	U	SMPL
J4272-02DUP - 1	JC-03-080918-ADUP	132	0.0488	-7470A	-7470A	PPB	13 Aug 2018 10:18:28	U	SMPL
J4272-02MS - 1	JC-03-080918-AMS	13460	3.9011	-7470A	-7470A	PPB	13 Aug 2018 10:20:37	U	SMPL
J4272-02MSD - 1	JC-03-080918-AMSD	13862	4.0173	-7470A	-7470A	PPB	13 Aug 2018 10:22:44	U	SMPL
CCV60 - 1	CCV60	18035	5.2234	-7470A	-7470A	PPB	13 Aug 2018 10:24:50	U	SMPL
CCB60 - 1	CCB60	-31	0.0017	-7470A	-7470A	PPB	13 Aug 2018 10:26:57	U	SMPL
J4272-04 - 1	JC-03-080918-B	97	0.0387	-7470A	-7470A	PPB	13 Aug 2018 10:29:06	U	SMPL
J4272-06 - 1	JC-03-080918-C	132	0.0488	-7470A	-7470A	PPB	13 Aug 2018 10:31:13	U	SMPL
J4272-08 - 1	JC-03-080918-D	123	0.0462	-7470A	-7470A	PPB	13 Aug 2018 10:33:19	U	SMPL
J4272-10 - 1	JC-03-080918-E	123	0.0462	-7470A	-7470A	PPB	13 Aug 2018 10:35:26	U	SMPL
J4272-12 - 1	JC-03-080918-F	158	0.0564	-7470A	-7470A	PPB	13 Aug 2018 10:37:33	U	SMPL
J4272-14 - 1	JC-03-080918-G	170	0.0598	-7470A	-7470A	PPB	13 Aug 2018 10:39:41	U	SMPL
J4272-16 - 1	JC-03-080918-H	124	0.0465	-7470A	-7470A	PPB	13 Aug 2018 10:41:49	U	SMPL
J4272-18 - 1	JC-03-080918-I	119	0.0451	-7470A	-7470A	PPB	13 Aug 2018 10:43:57	U	SMPL
J4272-20 - 1	JC-03-080918-J	117	0.0445	-7470A	-7470A	PPB	13 Aug 2018 10:46:06	U	SMPL
J4272-22 - 1	JC-03-080918-K	114	0.0436	-7470A	-7470A	PPB	13 Aug 2018 10:48:13	U	SMPL
CCV61 - 1	CCV61	18043	5.2257	-7470A	-7470A	PPB	13 Aug 2018 10:50:19	U	SMPL
CCB61 - 1	CCB61	-64	-0.0078	-7470A	-7470A	PPB	13 Aug 2018 10:52:26	U	SMPL
J4274-02 - 1	HD-01-080918	98	0.0390	-7470A	-7470A	PPB	13 Aug 2018 10:54:34	U	SMPL
J4282-02 - 1	SU-02-080918-A	108	0.0419	-7470A	-7470A	PPB	13 Aug 2018 10:56:41	U	SMPL
J4282-04 - 1	SU-02-080918-B	110	0.0425	-7470A	-7470A	PPB	13 Aug 2018 10:58:48	U	SMPL
J4405-04 - 1	EXCAVATION-SOIL	101	0.0399	-7470A	-7470A	PPB	13 Aug 2018 11:00:55	U	SMPL
J4417-01 - 1	SOIL/CUTTINGS	129	0.0480	-7470A	-7470A	PPB	13 Aug 2018 11:03:02	U	SMPL
PB111955BL - 1	PBW	120	0.0454	-7470A	-7470A	PPB	13 Aug 2018 11:05:10	U	SMPL
PB111955BS - 1	LCSW	13433	3.8933	-7470A	-7470A	PPB	13 Aug 2018 11:07:18	U	SMPL
PB111955TB - 1	PB111922TB	8	0.0130	-7470A	-7470A	PPB	13 Aug 2018 11:09:27	U	SMPL
J4409-02 - 1	2018-NYSEG-CLARK-AREA-9B	121	0.0457	-7470A	-7470A	PPB	13 Aug 2018 11:11:35	U	SMPL
J4409-02DUP - 1	2018-NYSEG-CLARK-AREA-9BDUP	129	0.0480	-7470A	-7470A	PPB	13 Aug 2018 11:13:42	U	SMPL
CCV62 - 1	CCV62	17799	5.1552	-7470A	-7470A	PPB	13 Aug 2018 11:15:50	U	SMPL
CCB62 - 1	CCB62	-200	-0.0471	-7470A	-7470A	PPB	13 Aug 2018 11:17:57	U	SMPL
J4409-02MS - 1	2018-NYSEG-CLARK-AREA-9BMS	14194	4.1132	-7470A	-7470A	PPB	13 Aug 2018 11:20:05	U	SMPL
J4409-02MSD - 1	2018-NYSEG-CLARK-AREA-9BMSD	13851	4.0141	-7470A	-7470A	PPB	13 Aug 2018 11:22:12	U	SMPL
J4409-04 - 1	2018-NYSEG-CLARK-AREA-10B	-21	0.0046	-7470A	-7470A	PPB	13 Aug 2018 11:24:20	U	SMPL
J4409-06 - 1	2018-NYSEG-CLARK-AREA-10C	84	0.0350	-7470A	-7470A	PPB	13 Aug 2018 11:26:27	U	SMPL
J4409-08 - 1	2018-NYSEG-CLARK-AREA-10D	125	0.0468	-7470A	-7470A	PPB	13 Aug 2018 11:28:35	U	SMPL
J4409-10 - 1	2018-NYSEG-CLARK-AREA-11B	145	0.0526	-7470A	-7470A	PPB	13 Aug 2018 11:30:42	U	SMPL
J4409-12 - 1	2018-NYSEG-CLARK-AREA-12B	113	0.0434	-7470A	-7470A	PPB	13 Aug 2018 11:32:51	U	SMPL
J4409-14 - 1	2018-NYSEG-CLARK-AREA-12C	115	0.0439	-7470A	-7470A	PPB	13 Aug 2018 11:34:59	U	SMPL
PB111956BL - 1	PBW	106	0.0413	-7470A	-7470A	PPB	13 Aug 2018 11:37:07	U	SMPL
PB111956BS - 1	LCSW	13559	3.9297	-7470A	-7470A	PPB	13 Aug 2018 11:39:14	U	SMPL
CCV63 - 1	CCV63	17464	5.0584	-7470A	-7470A	PPB	13 Aug 2018 11:41:21	U	SMPL
CCB63 - 1	CCB63	-162	-0.0361	-7470A	-7470A	PPB	13 Aug 2018 11:43:29	U	SMPL
J4379-19 - 1	MW-102S	-98	-0.0176	-7470A	-7470A	PPB	13 Aug 2018 11:45:37	U	SMPL
J4379-19DUP - 1	MW-102SDUP	91	0.0370	-7470A	-7470A	PPB	13 Aug 2018 11:47:44	U	SMPL
J4379-19MS - 1	MW-102SMS	12981	3.7626	-7470A	-7470A	PPB	13 Aug 2018 11:49:51	U	SMPL
J4379-19MSD - 1	MW-102SMSD	14242	4.1271	-7470A	-7470A	PPB	13 Aug 2018 11:51:59	U	SMPL
J4379-20 - 1	MW-102I	-48	-0.0032	-7470A	-7470A	PPB	13 Aug 2018 11:54:07	U	SMPL
J4379-21 - 1	MW-102D	187	0.0647	-7470A	-7470A	PPB	13 Aug 2018 11:56:15	U	SMPL
J4379-22 - 1	MW-101S	269	0.0884	-7470A	-7470A	PPB	13 Aug 2018 11:58:23	U	SMPL
J4379-23 - 1	MW-101I	119	0.0451	-7470A	-7470A	PPB	13 Aug 2018 12:00:31	U	SMPL
J4379-24 - 1	MW-101D	138	0.0506	-7470A	-7470A	PPB	13 Aug 2018 12:02:40	U	SMPL
J4379-25 - 1	MW-103D	104	0.0408	-7470A	-7470A	PPB	13 Aug 2018 12:04:47	U	SMPL
CCV64 - 1	CCV64	17533	5.0783	-7470A	-7470A	PPB	13 Aug 2018 12:06:54	U	SMPL
CCB64 - 1	CCB64	-89	-0.0150	-7470A	-7470A	PPB	13 Aug 2018 12:09:01	U	SMPL
J4272-02LX5 - 1		313	0.1012	-7470A	-7470A	PPB	13 Aug 2018 12:11:10	U	SMPL
J4272-02A - 1		14100	4.0861	-7470A	-7470A	PPB	13 Aug 2018 12:13:17	U	SMPL
J4409-02LX5 - 1		187	0.0647	-7470A	-7470A	PPB	13 Aug 2018 12:15:24	U	SMPL
J4409-02A - 1		14541	4.2135	-7470A	-7470A	PPB	13 Aug 2018 12:17:31	U	SMPL
J4379-19LX5 - 1		137	0.0503	-7470A	-7470A	PPB	13 Aug 2018 12:19:39	U	SMPL
J4379-19A - 1		14376	4.1658	-7470A	-7470A	PPB	13 Aug 2018 12:21:47	U	SMPL
CCV65 - 1	CCV65	18162	5.2601	-7470A	-7470A	PPB	13 Aug 2018 12:25:53	U	SMPL
CCB65 - 1	CCB65	17	0.0156	-7470A	-7470A	PPB	13 Aug 2018 12:58:01	U	SMPL