

245.1


LB 97257

INSTRUMENT ID: CX1

Linear

 μ Abs.:

33727



A= 0.0000e+000

B= 2.9750e-004

C= 1.1827e-002

Rho= 0.9999416

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.002	0.002	-34	0.000	-34				
0.05	0.050	0.065	0.015	178	0.0 %	178				
0.2	0.200	0.204	0.004	647	0.0 %	647				
2.5	2.500	2.474	-0.026	8275	0.0 %	8275				
5.0	5.000	5.041	0.041	16905	0.0 %	16905				
7.5	7.500	7.419	-0.081	24898	0.0 %	24898				
10.0	10.000	10.046	0.046	33727	0.0 %	33727				

LB97257 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 13 Aug 2018 14:12:24

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-34	-	0.0000	245.1	PPB	13 Aug 2018 14:16:09	S	Std
0.05 - 1	50.05	178	-	0.0500	245.1	PPB	13 Aug 2018 14:18:17	S	Std
0.2 - 1	50.2	647	-	0.2000	245.1	PPB	13 Aug 2018 14:20:25	S	Std
2.5 - 1	52.5	8275	-	2.5000	245.1	PPB	13 Aug 2018 14:22:33	S	Std
5.0 - 1	55	16905	-	5.0000	245.1	PPB	13 Aug 2018 14:24:41	S	Std
7.5 - 1	57.5	24898	-	7.5000	245.1	PPB	13 Aug 2018 14:26:48	S	Std
10.0 - 1	510	33727	-	10.0000	245.1	PPB	13 Aug 2018 14:31:25	S	Std
ICV04 - 1	ICV04	13098	3.9085	-245.1		PPB	13 Aug 2018 14:34:05	U	SMPL
ICB04 - 1	ICB04	-229	-0.0563	-245.1		PPB	13 Aug 2018 14:36:12	U	SMPL
CCV66 - 1	CCV66	17020	5.0753	-245.1		PPB	13 Aug 2018 14:38:20	U	SMPL
CCB66 - 1	CCB66	-242	-0.0602	-245.1		PPB	13 Aug 2018 14:40:26	U	SMPL
CRA - 1	CRA	127	0.0496	-245.1		PPB	13 Aug 2018 14:42:35	U	SMPL
HighStd - 1	HighStd	33385	9.9438	-245.1		PPB	13 Aug 2018 14:44:41	U	SMPL
ChkStd - 1	ChkStd	24093	7.1795	-245.1		PPB	13 Aug 2018 14:46:48	U	SMPL
PB111957BL - 1	PBW	-158	-0.0352	-245.1		PPB	13 Aug 2018 14:48:57	U	SMPL
PB111957BS - 1	LCSW	13962	4.1655	-245.1		PPB	13 Aug 2018 14:51:06	U	SMPL
J4412-01 - 1	080718-TIPC-E-D-M	-38	0.0005	-245.1		PPB	13 Aug 2018 14:53:14	U	SMPL
J4412-02 - 1	080718-TIPC-E-U-M	111	0.0448	-245.1		PPB	13 Aug 2018 14:55:23	U	SMPL
J4413-01 - 1	080818-TIPC-E-D-M	128	0.0499	-245.1		PPB	13 Aug 2018 14:57:32	U	SMPL
J4413-02 - 1	080818-TIPC-E-U-M	118	0.0469	-245.1		PPB	13 Aug 2018 14:59:40	U	SMPL
J4413-02DUP - 1	080818-TIPC-E-U-MDUP	5	0.0133	-245.1		PPB	13 Aug 2018 15:04:10	U	SMPL
CCV67 - 1	CCV67	17157	5.1160	-245.1		PPB	13 Aug 2018 15:06:16	U	SMPL
CCB67 - 1	CCB67	-212	-0.0512	-245.1		PPB	13 Aug 2018 15:08:23	U	SMPL
J4413-02MS - 1	080818-TIPC-E-U-MMS	15388	4.5897	-245.1		PPB	13 Aug 2018 15:10:31	U	SMPL
J4413-02MSD - 1	080818-TIPC-E-U-MMSD	15089	4.5008	-245.1		PPB	13 Aug 2018 15:12:38	U	SMPL
J4413-02LX5 - 1		-175	-0.0402	-245.1		PPB	13 Aug 2018 15:14:45	U	SMPL
J4413-02A - 1		13756	4.1042	-245.1		PPB	13 Aug 2018 15:16:52	U	SMPL
CCV68 - 1	CCV68	17422	5.1949	-245.1		PPB	13 Aug 2018 15:18:59	U	SMPL
CCB68 - 1	CCB68	-235	-0.0581	-245.1		PPB	13 Aug 2018 15:21:07	U	SMPL