

LB97260


7471B

INSTRUMENT ID: CY2

Linear

μ Abs.:

45288



A= 0.0000e+000

B= 2.1412e-004

C= 2.7048e-002

Rho= 0.9987097

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.035	0.035	35	0.000	35				
0.2	0.200	0.157	-0.043	605	0.0 %	605				
2.5	2.500	2.386	-0.114	11015	0.0 %	11015				
5.0	5.000	5.060	0.060	23507	0.0 %	23507				
7.5	7.500	7.839	0.339	36483	0.0 %	36483				
10.0	10.000	9.724	-0.276	45288	0.0 %	45288				

LB97260 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 13 Aug 2018 15:47:49

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc/Method	Units	Date	Type	Type
0.0 - 1	50	35	-	0.00007471B	PPB	13 Aug 2018 15:50:15	S	Std
0.2 - 1	50.2	605	-	0.20007471B	PPB	13 Aug 2018 15:52:33	S	Std
2.5 - 1	52.5	11015	-	2.50007471B	PPB	13 Aug 2018 15:54:50	S	Std
5.0 - 1	55	23507	-	5.00007471B	PPB	13 Aug 2018 15:57:08	S	Std
7.5 - 1	57.8	36483	-	7.50007471B	PPB	13 Aug 2018 15:59:25	S	Std
10.0 - 1	510	45288	-	10.00007471B	PPB	13 Aug 2018 16:01:42	S	Std
ICV05 - 1	ICV05	18511	3.9906	-7471B	PPB	13 Aug 2018 16:05:00	U	SMPL
ICB05 - 1	ICB05	-154	-0.0059	-7471B	PPB	13 Aug 2018 16:07:16	U	SMPL
CCV69 - 1	CCV69	23379	5.0330	-7471B	PPB	13 Aug 2018 16:09:32	U	SMPL
CCB69 - 1	CCB69	-156	-0.0064	-7471B	PPB	13 Aug 2018 16:11:48	U	SMPL
CRA - 1	CRA	828	0.2043	-7471B	PPB	13 Aug 2018 16:14:04	U	SMPL
HighStd - 1	HighStd	44615	9.5800	-7471B	PPB	13 Aug 2018 16:16:20	U	SMPL
ChkStd - 1	ChkStd	34319	7.3754	-7471B	PPB	13 Aug 2018 16:18:36	U	SMPL
PB111968BL - 1	PBS	-232	-0.0226	-7471B	PPB	13 Aug 2018 16:20:54	U	SMPL
PB111968BS - 1	LCSS	17166	3.7026	-7471B	PPB	13 Aug 2018 16:23:13	U	SMPL
J4409-11DLX5 - 1	2018-NYSEG-CLARK-AREA-12B	33112	7.1170	-7471B	PPB	13 Aug 2018 16:25:29	U	SMPL
J4423-01 - 1	72-11944	44163	9.4832	-7471B	PPB	13 Aug 2018 16:27:47	U	SMPL
J4427-01 - 1	RB47197	86	0.0455	-7471B	PPB	13 Aug 2018 16:33:25	U	SMPL
J4430-12 - 1	G.L-WC-01	18998	4.0949	-7471B	PPB	13 Aug 2018 16:35:43	U	SMPL
J4442-01 - 1	SP-1-1	1056	0.2532	-7471B	PPB	13 Aug 2018 16:37:59	U	SMPL
CCV70 - 1	CCV70	22978	4.9471	-7471B	PPB	13 Aug 2018 16:40:16	U	SMPL
CCB70 - 1	CCB70	-172	-0.0098	-7471B	PPB	13 Aug 2018 16:42:31	U	SMPL
J4442-11 - 1	SP-1-2	1645	0.3793	-7471B	PPB	13 Aug 2018 16:44:48	U	SMPL
J4442-21 - 1	SP-1-3	339	0.0996	-7471B	PPB	13 Aug 2018 16:47:04	U	SMPL
J4442-31 - 1	SP-1-4	2430	0.5474	-7471B	PPB	13 Aug 2018 16:49:20	U	SMPL
J4442-41 - 1	SP-1-5	842	0.2073	-7471B	PPB	13 Aug 2018 16:51:36	U	SMPL
J4442-51 - 1	SP-1-6	256	0.0819	-7471B	PPB	13 Aug 2018 16:53:52	U	SMPL
J4443-01 - 1	SALEM-RD-SP	3657	0.8101	-7471B	PPB	13 Aug 2018 16:56:09	U	SMPL
J4447-06 - 1	SP-1-7	1049	0.2517	-7471B	PPB	13 Aug 2018 16:58:26	U	SMPL
J4447-16 - 1	SP-1-8	2483	0.5587	-7471B	PPB	13 Aug 2018 17:00:44	U	SMPL
J4447-26 - 1	SP-1-9	2155	0.4885	-7471B	PPB	13 Aug 2018 17:03:01	U	SMPL
J4447-26DUP - 1	SP-1-9DUP	2087	0.4739	-7471B	PPB	13 Aug 2018 17:05:18	U	SMPL
CCV71 - 1	CCV71	22945	4.9400	-7471B	PPB	13 Aug 2018 17:07:33	U	SMPL
CCB71 - 1	CCB71	-114	0.0026	-7471B	PPB	13 Aug 2018 17:09:49	U	SMPL
J4447-26MS - 1	SP-1-9MS	19125	4.1221	-7471B	PPB	13 Aug 2018 17:12:07	U	SMPL
J4447-26MSD - 1	SP-1-9MSD	18485	3.9851	-7471B	PPB	13 Aug 2018 17:14:23	U	SMPL
J4447-36 - 1	SP-1-10	1891	0.4319	-7471B	PPB	13 Aug 2018 17:16:40	U	SMPL
J4447-26LX5 - 1		380	0.1084	-7471B	PPB	13 Aug 2018 17:18:56	U	SMPL
J4447-26A - 1		20307	4.3752	-7471B	PPB	13 Aug 2018 17:21:13	U	SMPL
CCV72 - 1	CCV72	23009	4.9537	-7471B	PPB	13 Aug 2018 17:23:30	U	SMPL
CCB72 - 1	CCB72	-150	-0.0051	-7471B	PPB	13 Aug 2018 17:25:47	U	SMPL