



Analytical Summary Report

Analysis Method: 9040C

Analyst By : sweta

Parameter: pH

Supervisor Review By : apatel

Run Number: LB97271

Slope : 98.6

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			22.0	4.01	08/13/2018	17:10
2	CAL2	1	Water			21.9	7.01	08/13/2018	17:12
3	CAL3	1	Water			21.8	10.04	08/13/2018	17:14
4	ICV	1	Water			22.0	7.00	08/13/2018	17:16
5	CCV1	1	Water			21.9	2.00	08/13/2018	17:18
6	J4459-01	1	Water			24.0	7.06	08/13/2018	17:20
7	J4459-01DUP	1	Water			24.1	7.07	08/13/2018	17:22
8	CCV2	1	Water			21.8	12.00	08/13/2018	17:24

WORKLIST(Hardcopy Internal Chain)

LB97271

WorkList Name : pH081318-1

WorkList ID : 115571

Date : 8/13/2018 4:54:44 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/13/2018	Water	J4459-01	pH	Cool 4 deg C	PSEG01	P63	FRAC-TANK-SV30518L	08/13/2018	9040C

Date/Time 08/13/18 5:00
 Received by: SP
 Relinquished by: LB

Date/Time 08/13/18 5:40
 Received by: BP
 Relinquished by: SP