



Analytical Summary Report

Analysis Method: 9040C

Analyst By : Sweta

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB97273

Slope : 98.6

BalanceID: WC SC-4

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			22.0	4.01	08/13/2018	17:10
2	CAL2	1	Water			21.9	7.01	08/13/2018	17:12
3	CAL3	1	Water			21.8	10.04	08/13/2018	17:14
4	ICV	1	Water			22.0	7.00	08/13/2018	17:16
5	CCV1	1	Water			21.9	2.00	08/13/2018	17:30
6	J4454-01	1	Solid	20.02	20	23.4	6.12	08/13/2018	17:32
7	J4454-01DUP	1	Solid	20.04	20	23.5	6.13	08/13/2018	17:34
8	CCV2	1	Water			22.0	12.00	08/13/2018	17:36

WORKLIST(Hardcopy Internal Chain)

LB 97273

WorkList Name : corrosivity081318

WorkList ID : 115570

Date : 8/13/2018 4:54:36 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/10/2018	Solid	J4454-01	Corrosivity	Cool 4 deg C	USAR03	P12	CB-0818-S-TCLP-61A	08/10/2018	9040C

Date/Time 08/13/18 5:00
 Received by: SP
 Relinquished by: SP

Date/Time 08/13/18 5:40
 Received by: SP
 Relinquished by: SP