



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB97279

Ambient Barometric Pressure (mmHg): 755

Barometric Scale ID: 051106

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.0	80.6	80.9	08/14/2018	09:45
2	J4458-01		1	8	>100.0	>212.0	>212.0	08/14/2018	10:10
3	J4458-01DUP		1	8	>100.0	>212.0	>212.0	08/14/2018	10:31
4	J4458-02		1	9	>100.0	>212.0	>212.0	08/14/2018	10:50
5	J4458-03		1	8	>100.0	>212.0	>212.0	08/14/2018	11:15
6	J4459-02		1	7	>100.0	>212.0	>212.0	08/14/2018	11:50

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

LB 97279

WORKLIST(Hardcopy Internal Chain)

WorkList Name : fp81418

WorkList ID : 115605

Date : 8/14/2018 9:04:55 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/26/2018	Water	J4458-01	Flash Point	Cool 4 deg C	DIST01	P53	MH-92A	08/12/2018	1010A
08/26/2018	Water	J4458-02	Flash Point	Cool 4 deg C	DIST01	P53	MH-93	08/12/2018	1010A
08/26/2018	Water	J4458-03	Flash Point	Cool 4 deg C	DIST01	P53	MH-94	08/12/2018	1010A
08/16/2018	Water	J4459-02	Flash Point	Cool 4 deg C	PSEG01	P63	FRAC-TANK-SV30518L	08/13/2018	1010A

Date/Time

Received by:

Relinquished by:

08/14/18 9:15 AM

KS

SP

Date/Time

Received by:

Relinquished by:

08/14/18 11:55 AM

SP

KS