

LB 97283


7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

46150



A= 0.0000e+000

B= 2.1715e-004

C= 4.8045e-002

Rho= 0.9998062

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.052	0.052	16	0.000	16				
0.2	0.200	0.240	0.040	883	0.0 %	883				
2.5	2.500	2.480	-0.020	11200	0.0 %	11200				
5.0	5.000	4.851	-0.149	22119	0.0 %	22119				
7.5	7.500	7.508	0.008	34351	0.0 %	34351				
10.0	10.000	10.070	0.070	46150	0.0 %	46150				

LB97283 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 14 Aug 2018 15:06:40

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	16	-	0.0000	7471B	PPB	14 Aug 2018 15:09:31	S	Std
0.2 - 1	50.2	883	-	0.2000	7471B	PPB	14 Aug 2018 15:11:48	S	Std
2.5 - 1	52.5	11200	-	2.5000	7471B	PPB	14 Aug 2018 15:14:06	S	Std
5.0 - 1	55	22119	-	5.0000	7471B	PPB	14 Aug 2018 15:16:23	S	Std
7.5 - 1	57.5	34351	-	7.5000	7471B	PPB	14 Aug 2018 15:18:40	S	Std
10.0 - 1	510	46150	-	10.0000	7471B	PPB	14 Aug 2018 15:20:58	S	Std
ICV07 - 1	ICV07	18029	3.9631	-	7471B	PPB	14 Aug 2018 15:24:12	U	SMPL
ICB07 - 1	ICB07	-150	0.0155	-	7471B	PPB	14 Aug 2018 15:26:28	U	SMPL
CCV78 - 1	CCV78	22450	4.9232	-	7471B	PPB	14 Aug 2018 15:28:45	U	SMPL
CCB78 - 1	CCB78	-145	0.0166	-	7471B	PPB	14 Aug 2018 15:31:00	U	SMPL
CRA - 1	CRA	886	0.2404	-	7471B	PPB	14 Aug 2018 15:33:17	U	SMPL
HighStd - 1	HighStd	46475	10.1403	-	7471B	PPB	14 Aug 2018 15:35:33	U	SMPL
ChkStd - 1	ChkStd	33143	7.2452	-	7471B	PPB	14 Aug 2018 15:37:49	U	SMPL
PB112008BL - 1	PBS	-300	-0.0171	-	7471B	PPB	14 Aug 2018 15:40:07	U	SMPL
PB112008BS - 1	LCSS	15745	3.4671	-	7471B	PPB	14 Aug 2018 15:42:26	U	SMPL
J4449-01 - 1	TILCON-MH-001	377	0.1299	-	7471B	PPB	14 Aug 2018 15:44:42	U	SMPL
J4449-02 - 1	TILCON-PL-001	584	0.1749	-	7471B	PPB	14 Aug 2018 15:47:00	U	SMPL
J4460-01 - 1	RA-080918-FL-001	4440	1.0122	-	7471B	PPB	14 Aug 2018 15:49:17	U	SMPL
J4460-02 - 1	RA-080918-SW-002	6643	1.4906	-	7471B	PPB	14 Aug 2018 15:51:35	U	SMPL
J4460-03 - 1	RA-081018-SW-003	2472	0.5849	-	7471B	PPB	14 Aug 2018 15:53:51	U	SMPL
CCV79 - 1	CCV79	22255	4.8808	-	7471B	PPB	14 Aug 2018 15:56:07	U	SMPL
CCB79 - 1	CCB79	-200	0.0046	-	7471B	PPB	14 Aug 2018 15:58:23	U	SMPL
J4460-09 - 1	RA-081018-FL-009	4158	0.9510	-	7471B	PPB	14 Aug 2018 16:00:40	U	SMPL
J4460-09DUP - 1	RA-081018-FL-009DUP	4236	0.9679	-	7471B	PPB	14 Aug 2018 16:02:56	U	SMPL
J4460-09MS - 1	RA-081018-FL-009MS	21165	4.6441	-	7471B	PPB	14 Aug 2018 16:05:12	U	SMPL
J4460-09MSD - 1	RA-081018-FL-009MSD	20620	4.5258	-	7471B	PPB	14 Aug 2018 16:07:28	U	SMPL
J4460-11 - 1	RA-081018-FL-011	4695	1.0676	-	7471B	PPB	14 Aug 2018 16:09:45	U	SMPL
J4460-14 - 1	RA-081018-FL-014	1550	0.3846	-	7471B	PPB	14 Aug 2018 16:12:01	U	SMPL
J4460-15 - 1	RA-081018-ESW-015	7398	1.6546	-	7471B	PPB	14 Aug 2018 16:14:19	U	SMPL
J4460-09LX5 - 1		515	0.1599	-	7471B	PPB	14 Aug 2018 16:16:37	U	SMPL
J4460-09A - 1		21518	4.7208	-	7471B	PPB	14 Aug 2018 16:18:54	U	SMPL
CCV80 - 1	CCV80	22308	4.8923	-	7471B	PPB	14 Aug 2018 16:21:11	U	SMPL
CCB80 - 1	CCB80	-177	0.0096	-	7471B	PPB	14 Aug 2018 16:23:28	U	SMPL