


ISM02.4

LB 97286
INSTRUMENT ID: CV4

Linear

 μ Abs.:

4689E



A= 0.0000e+000

B= 2.1361e-004

C= -5.0153e-003

Rho= 0.9999944

Accept=Accepted

slope
7.1515e-4

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.002	0.002	33	0.000	33					-
0.2	0.200	0.194	-0.006	930	0.0 %	930					-3
2.5	2.500	2.518	0.018	11812	0.0 %	11812					1
5.0	5.000	4.989	-0.011	23381	0.0 %	23381					0
7.5	7.500	7.484	-0.016	35058	0.0 %	35058					0
10.0	10.000	10.013	0.013	46898	0.0 %	46898					0

hb
8/14/18

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0.0	33	-		0 ISM02.4	PPB	8/14/2018 16:44	Std
	0.2 S0.2	930	-		0.2 ISM02.4	PPB	8/14/2018 16:47	Std
	2.5 S2.5	11812	-		2.5 ISM02.4	PPB	8/14/2018 16:51	Std
	5 S5	23381	-		5 ISM02.4	PPB	8/14/2018 16:53	Std
	7.5 S7.5	35058	-		7.5 ISM02.4	PPB	8/14/2018 16:55	Std
	10 S10.0	46898	-		10 ISM02.4	PPB	8/14/2018 16:58	Std
ICV	ICV	18035	3.8475 -		ISM02.4	PPB	8/14/2018 17:00	SMPL
ICB	ICB	-153	-0.0377 -		ISM02.4	PPB	8/14/2018 17:03	SMPL
CCV081	CCV081	23621	5.0407 -		ISM02.4	PPB	8/14/2018 17:05	SMPL
CCB081	CCB081	-159	-0.039 -		ISM02.4	PPB	8/14/2018 17:07	SMPL
PB112024BL	PBS004	2	-0.0046 -		ISM02.4	PPB	8/14/2018 17:10	SMPL
J4425-03	MA3YQ8	-15	-0.0082 -		ISM02.4	PPB	8/14/2018 17:14	SMPL
J4425-04	MA3YR4	1434	0.3013 -		ISM02.4	PPB	8/14/2018 17:16	SMPL
J4425-05	MA3YR7	167	0.0307 -		ISM02.4	PPB	8/14/2018 17:19	SMPL
J4425-06	MA3YR7D	169	0.0311 -		ISM02.4	PPB	8/14/2018 17:21	SMPL
J4425-07	MA3YR7S	12440	2.6523 -		ISM02.4	PPB	8/14/2018 17:23	SMPL
J4425-09	MA3YS7	3732	0.7922 -		ISM02.4	PPB	8/14/2018 17:25	SMPL
J4425-10	MA3YS9	584	0.1197 -		ISM02.4	PPB	8/14/2018 17:28	SMPL
J4425-11	MA3Z58	20994	4.4795 -		ISM02.4	PPB	8/14/2018 17:30	SMPL
J4425-12	MA3Z60	5018	1.0669 -		ISM02.4	PPB	8/14/2018 17:32	SMPL
J4425-13	MA3Z61	2217	0.4686 -		ISM02.4	PPB	8/14/2018 17:35	SMPL
J4425-14	MA3Z63	1576	0.3316 -		ISM02.4	PPB	8/14/2018 17:37	SMPL
J4425-15	MA3YS4	5492	1.1681 -		ISM02.4	PPB	8/14/2018 17:39	SMPL
J4425-01	MA3YQ4	10278	2.1905 -		ISM02.4	PPB	8/14/2018 17:41	SMPL
CCV082	CCV082	22708	4.8457 -		ISM02.4	PPB	8/14/2018 17:51	SMPL
CCB082	CCB082	-125	-0.0317 -		ISM02.4	PPB	8/14/2018 17:53	SMPL