



Analytical Summary Report

Analysis Method: SM 4500-PH B

Analyst By : sweta

Parameter: pH

Supervisor Review By : apatel

Run Number: LB97290

Slope : 98.8

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.8	4.01	08/14/2018	12:40
2	CAL2	1	Water			21.9	7.00	08/14/2018	12:42
3	CAL3	1	Water			21.9	10.04	08/14/2018	12:44
4	ICV	1	Water			22.0	7.00	08/14/2018	12:46
5	CCV1	1	Water			21.7	2.00	08/14/2018	12:48
6	J4470-01	1	Water			18.9	5.20	08/14/2018	12:50
7	J4470-01DUP	1	Water			19.0	5.21	08/14/2018	12:52
8	CCV2	1	Water			22.0	12.00	08/14/2018	12:54

WORKLIST(Hardcopy Internal Chain)

LB9A290

WorkList Name : pH081418

WorkList ID : 115617

Date : 8/14/2018 12:27:43 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/13/2018	Water	J4470-01 B	pH	Cool 4 deg C	STER12	P31	CELL-POOL-WATER-3RD-Q	08/13/2018	SM 4500-PH B

Date/Time 08/14/18 12:30
 Received by: SP
 Relinquished by: SP

Date/Time 08/14/18 12:10
 Received by: SP
 Relinquished by: SP