



Analytical Summary Report

Analysis Method: 9040C

Analyst By : sweta

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB97292

Slope : 99.0

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.9	4.01	08/14/2018	16:20
2	CAL2	1	Water			21.8	7.01	08/14/2018	16:22
3	CAL3	1	Water			22.0	10.04	08/14/2018	16:24
4	ICV	1	Water			21.9	7.00	08/14/2018	16:26
5	CCV1	1	Water			21.7	2.00	08/14/2018	16:28
6	J4474-02	1	Water			25.6	6.93	08/14/2018	16:30
7	J4474-02DUP	1	Water			25.7	6.94	08/14/2018	16:32
8	CCV2	1	Water			22.0	12.00	08/14/2018	16:34

UBA 7292

WORKLIST(Hardcopy Internal Chain)

WorkList Name : corrosivity081418-1

WorkList ID : 115626

Date : 8/14/2018 3:52:19 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/14/2018	Water	J4474-02	Corrosivity	Cool 4 deg C	PSEG01	K63	FRAC-TANK-SV28570L	08/14/2018	9040C

Date/Time 08/14/18 4:10
Received by: SP
Relinquished by: EJP

Date/Time 08/14/18 4:40
Received by: SP
Relinquished by: SP