



Analytical Summary Report

Analysis Method: 9040C

Analyst By : Sweta

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB97293

Slope : 99.0

BalanceID: WC SC-4

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.9	4.01	08/14/2018	16:20
2	CAL2	1	Water			21.8	7.01	08/14/2018	16:22
3	CAL3	1	Water			22.0	10.04	08/14/2018	16:24
4	ICV	1	Water			21.9	7.00	08/14/2018	16:26
5	CCV1	1	Water			21.7	2.00	08/14/2018	16:45
6	J4471-04	1	Solid	20.02	20	24.0	6.04	08/14/2018	16:47
7	J4471-04DUP	1	Solid	20.03	20	24.1	6.05	08/14/2018	16:49
8	J4471-08	1	Solid	20.02	20	23.9	8.47	08/14/2018	16:51
9	J4471-12	1	Solid	20.02	20	23.5	8.64	08/14/2018	16:53
10	J4471-16	1	Solid	20.04	20	23.3	8.06	08/14/2018	16:55
11	J4471-20	1	Solid	20.03	20	23.0	8.67	08/14/2018	16:57
12	J4471-24	1	Solid	20.03	20	22.9	8.82	08/14/2018	16:59
13	J4471-27	1	Solid	20.02	20	22.7	8.20	08/14/2018	17:01
14	CCV2	1	Water			22.0	12.00	08/14/2018	17:03

WORKLIST(Hardcopy Internal Chain)

CB97293

WorkList Name : corrosivity081418

WorkList ID : 115625

Date : 8/14/2018 3:52:07 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/14/2018	Solid	J4471-04	Corrosivity	Cool 4 deg C	PSEG01	K63	TP-1	08/14/2018	9040C
08/14/2018	Solid	J4471-08	Corrosivity	Cool 4 deg C	PSEG01	K63	TP-2	08/14/2018	9040C
08/14/2018	Solid	J4471-12	Corrosivity	Cool 4 deg C	PSEG01	K63	TP-3	08/14/2018	9040C
08/14/2018	Solid	J4471-16	Corrosivity	Cool 4 deg C	PSEG01	K63	TP-4	08/14/2018	9040C
08/14/2018	Solid	J4471-20	Corrosivity	Cool 4 deg C	PSEG01	K63	TP-5	08/14/2018	9040C
08/14/2018	Solid	J4471-24	Corrosivity	Cool 4 deg C	PSEG01	K63	TP-6	08/14/2018	9040C
08/14/2018	Solid	J4471-27	Corrosivity	Cool 4 deg C	PSEG01	K63	BED-ROCK	08/14/2018	9040C

Date/Time 08/14/18 4:10
 Received by: SP
 Relinquished by: SP

Date/Time 08/14/18 5:10
 Received by: SP
 Relinquished by: SP