


7470A

L397298
INSTRUMENT ID: CV1

 μ Abs.:

3455E



Linear

A= 0.0000e+000

B= 2.8400e-004

C= 3.1843e-002

Rho= 0.9994491

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.029	0.029	-11	0.000	-11				
0.2	0.200	0.209	0.009	624	0.0 %	624				
2.5	2.500	2.378	-0.122	8260	0.0 %	8260				
5.0	5.000	5.014	0.014	17544	0.0 %	17544				
7.5	7.500	7.724	0.224	27086	0.0 %	27086				
10.0	10.000	9.846	-0.154	34558	0.0 %	34558				

LB97298 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 15 Aug 2018 08:44:35

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-11	-	0.00007470A	PPB		15 Aug 2018 08:50:55	S	Std
0.2 - 1	50.2	624	-	0.20007470A	PPB		15 Aug 2018 08:53:04	S	Std
2.5 - 1	52.5	8260	-	2.50007470A	PPB		15 Aug 2018 08:55:12	S	Std
5.0 - 1	55	17544	-	5.00007470A	PPB		15 Aug 2018 08:57:21	S	Std
7.5 - 1	57.5	27086	-	7.50007470A	PPB		15 Aug 2018 08:59:28	S	Std
10.0 - 1	510	34558	-	10.00007470A	PPB		15 Aug 2018 09:04:03	S	Std
ICV08 - 1	ICV08	14041	4.0194	-7470A	PPB		15 Aug 2018 09:06:37	U	SMPL
ICB08 - 1	ICB08	-232	-0.0340	-7470A	PPB		15 Aug 2018 09:08:44	U	SMPL
CCV83 - 1	CCV83	18530	5.2943	-7470A	PPB		15 Aug 2018 09:10:51	U	SMPL
CCB83 - 1	CCB83	-232	-0.0340	-7470A	PPB		15 Aug 2018 09:12:58	U	SMPL
CRA - 1	CRA	616	0.2068	-7470A	PPB		15 Aug 2018 09:15:06	U	SMPL
HighStd - 1	HighStd	34944	9.9558	-7470A	PPB		15 Aug 2018 09:17:12	U	SMPL
ChkStd - 1	ChkStd	25418	7.2504	-7470A	PPB		15 Aug 2018 09:19:19	U	SMPL
PB112043BL - 1	PBW	-317	-0.0582	-7470A	PPB		15 Aug 2018 09:21:28	U	SMPL
PB112043BS - 1	LCSW	13042	3.7357	-7470A	PPB		15 Aug 2018 09:23:38	U	SMPL
PB112043TB - 1	PB112003TB	-159	-0.0133	-7470A	PPB		15 Aug 2018 09:25:45	U	SMPL
J4454-01 - 1	CB-0818-S-TCLP-61A	-26	0.0245	-7470A	PPB		15 Aug 2018 09:27:55	U	SMPL
J4454-01DUP - 1	CB-0818-S-TCLP-61ADUP	-24	0.0250	-7470A	PPB		15 Aug 2018 09:30:03	U	SMPL
J4454-01MS - 1	CB-0818-S-TCLP-61AMS	13906	3.9811	-7470A	PPB		15 Aug 2018 09:32:11	U	SMPL
J4454-01MSD - 1	CB-0818-S-TCLP-61AMSD	12927	3.7031	-7470A	PPB		15 Aug 2018 09:34:18	U	SMPL
CCV84 - 1	CCV84	18165	5.1906	-7470A	PPB		15 Aug 2018 09:36:25	U	SMPL
CCB84 - 1	CCB84	-218	-0.0301	-7470A	PPB		15 Aug 2018 09:38:32	U	SMPL
PB112046BL - 1	PBW	-49	0.0179	-7470A	PPB		15 Aug 2018 09:40:40	U	SMPL
PB112046BS - 1	LCSW	13954	3.9947	-7470A	PPB		15 Aug 2018 09:42:47	U	SMPL
PB112046TB - 1	PB112006TB	-163	-0.0144	-7470A	PPB		15 Aug 2018 09:44:53	U	SMPL
J4459-02 - 1	FRAC-TANK-SV30518L	-26	0.0245	-7470A	PPB		15 Aug 2018 09:47:01	U	SMPL
J4459-02DUP - 1	FRAC-TANK-SV30518LDUP	10	0.0347	-7470A	PPB		15 Aug 2018 09:49:08	U	SMPL
J4459-02MS - 1	FRAC-TANK-SV30518LMS	12653	3.6252	-7470A	PPB		15 Aug 2018 09:51:16	U	SMPL
J4459-02MSD - 1	FRAC-TANK-SV30518LMSD	13239	3.7917	-7470A	PPB		15 Aug 2018 09:53:23	U	SMPL
PB112048BL - 1	PBW	-172	-0.0170	-7470A	PPB		15 Aug 2018 09:55:32	U	SMPL
PB112048BS - 1	LCSW	13069	3.7434	-7470A	PPB		15 Aug 2018 09:57:42	U	SMPL
J4459-01 - 1	FRAC-TANK-SV30518L	418	0.1506	-7470A	PPB		15 Aug 2018 09:59:49	U	SMPL
CCV85 - 1	CCV85	18216	5.2051	-7470A	PPB		15 Aug 2018 10:01:57	U	SMPL
CCB85 - 1	CCB85	-208	-0.0272	-7470A	PPB		15 Aug 2018 10:04:04	U	SMPL
J4459-01DUP - 1	FRAC-TANK-SV30518LDUP	530	0.1824	-7470A	PPB		15 Aug 2018 10:06:12	U	SMPL
J4459-01MS - 1	FRAC-TANK-SV30518LMS	13119	3.7576	-7470A	PPB		15 Aug 2018 10:18:22	U	SMPL
J4459-01MSD - 1	FRAC-TANK-SV30518LMSD	13143	3.7644	-7470A	PPB		15 Aug 2018 10:20:28	U	SMPL
J4454-01LX5 - 1		-156	-0.0125	-7470A	PPB		15 Aug 2018 10:22:36	U	SMPL
J4454-01A - 1		13009	3.7263	-7470A	PPB		15 Aug 2018 10:24:43	U	SMPL
J4459-02LX5 - 1		-125	-0.0037	-7470A	PPB		15 Aug 2018 10:26:51	U	SMPL
J4459-02A - 1		13295	3.8076	-7470A	PPB		15 Aug 2018 10:28:59	U	SMPL
J4459-01LX5 - 1		-63	0.0140	-7470A	PPB		15 Aug 2018 10:31:07	U	SMPL
J4459-01A - 1		15451	4.4199	-7470A	PPB		15 Aug 2018 10:33:16	U	SMPL
CCV86 - 1	CCV86	18132	5.1813	-7470A	PPB		15 Aug 2018 10:35:23	U	SMPL
CCB86 - 1	CCB86	-199	-0.0247	-7470A	PPB		15 Aug 2018 10:37:31	U	SMPL