

245.1


LB 97299

INSTRUMENT ID: CVI

Linear

 μ Abs.:

34086



A= 0.0000e+000

B= 2.9245e-004

C= 1.4155e-002

Rho= 0.9998859

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.014	0.014	-2	0.000	-2				
0.05	0.050	0.049	-0.001	119	0.0 %	119				
0.2	0.200	0.199	-0.001	633	0.0 %	633				
2.5	2.500	2.535	0.035	8620	0.0 %	8620				
5.0	5.000	4.886	-0.114	16658	0.0 %	16658				
7.5	7.500	7.585	0.085	25888	0.0 %	25888				
10.0	10.000	9.982	-0.018	34086	0.0 %	34086				

LB97299 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 15 Aug 2018 10:47:42

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-2	-	0.0000	245.1	PPB	15 Aug 2018 10:50:13	S	Std
0.05 - 1	50.05	119	-	0.0500	245.1	PPB	15 Aug 2018 10:52:21	S	Std
0.2 - 1	50.2	633	-	0.2000	245.1	PPB	15 Aug 2018 10:54:30	S	Std
2.5 - 1	52.5	8620	-	2.5000	245.1	PPB	15 Aug 2018 10:56:38	S	Std
5.0 - 1	55	16658	-	5.0000	245.1	PPB	15 Aug 2018 10:58:46	S	Std
7.5 - 1	57.5	25888	-	7.5000	245.1	PPB	15 Aug 2018 11:00:54	S	Std
10.0 - 1	510	34086	-	10.0000	245.1	PPB	15 Aug 2018 11:05:51	S	Std
ICV09 - 1	ICV09	13082	3.8399	-245.1		PPB	15 Aug 2018 11:08:34	U	SMPL
ICB09 - 1	ICB09	-196	-0.0432	-245.1		PPB	15 Aug 2018 11:10:40	U	SMPL
CCV87 - 1	CCV87	17001	4.9860	-245.1		PPB	15 Aug 2018 11:12:48	U	SMPL
CCB87 - 1	CCB87	-201	-0.0446	-245.1		PPB	15 Aug 2018 11:14:55	U	SMPL
CRA - 1	CRA	107	0.0454	-245.1		PPB	15 Aug 2018 11:17:04	U	SMPL
HighStd - 1	HighStd	34576	10.1258	-245.1		PPB	15 Aug 2018 11:19:10	U	SMPL
ChkStd - 1	ChkStd	24332	7.1299	-245.1		PPB	15 Aug 2018 11:21:17	U	SMPL
PB112029BL - 1	PBW	-52	-0.0011	-245.1		PPB	15 Aug 2018 11:26:45	U	SMPL
PB112029BS - 1	LCSW	13052	3.8312	-245.1		PPB	15 Aug 2018 11:28:52	U	SMPL
J4413-01 - 1	080818-TIPC-E-D-M	-34	0.0042	-245.1		PPB	15 Aug 2018 11:33:31	U	SMPL
J4440-01 - 1	080918-TIPC-F-D-B	-31	0.0051	-245.1		PPB	15 Aug 2018 11:35:39	U	SMPL
J4440-02 - 1	080918-TIPC-F-D-S	-16	0.0095	-245.1		PPB	15 Aug 2018 11:37:47	U	SMPL
J4440-03 - 1	080918-TIPC-F-U-B	-24	0.0071	-245.1		PPB	15 Aug 2018 11:39:55	U	SMPL
J4440-04 - 1	080918-TIPC-F-U-S	-37	0.0033	-245.1		PPB	15 Aug 2018 11:42:02	U	SMPL
CCV88 - 1	CCV88	17004	4.9869	-245.1		PPB	15 Aug 2018 11:44:09	U	SMPL
CCB88 - 1	CCB88	-218	-0.0496	-245.1		PPB	15 Aug 2018 11:46:16	U	SMPL
J4458-01 - 1	MH-92A	325	0.1092	-245.1		PPB	15 Aug 2018 11:53:42	U	SMPL
J4458-01DUP - 1	MH-92ADUP	339	0.1133	-245.1		PPB	15 Aug 2018 12:01:20	U	SMPL
J4458-01MS - 1	MH-92AMS	15931	4.6731	-245.1		PPB	15 Aug 2018 12:03:26	U	SMPL
J4458-01MSD - 1	MH-92AMSD	15836	4.6453	-245.1		PPB	15 Aug 2018 12:05:33	U	SMPL
J4458-02 - 1	MH-93	9159	2.6927	-245.1		PPB	15 Aug 2018 12:07:41	U	SMPL
J4458-03 - 1	MH-94	6844	2.0157	-245.1		PPB	15 Aug 2018 12:09:49	U	SMPL
J4462-01 - 1	091318-TIPC-F-D-B	-111	-0.0183	-245.1		PPB	15 Aug 2018 12:11:58	U	SMPL
J4462-02 - 1	091318-TIPC-F-D-S	-48	0.0001	-245.1		PPB	15 Aug 2018 12:14:06	U	SMPL
J4462-03 - 1	091318-TIPC-F-U-B	-32	0.0048	-245.1		PPB	15 Aug 2018 12:16:14	U	SMPL
J4462-04 - 1	091318-TIPC-F-U-S	-26	0.0066	-245.1		PPB	15 Aug 2018 12:18:21	U	SMPL
CCV89 - 1	CCV89	16735	4.9082	-245.1		PPB	15 Aug 2018 12:20:28	U	SMPL
CCB89 - 1	CCB89	-146	-0.0285	-245.1		PPB	15 Aug 2018 12:22:35	U	SMPL
J4462-05 - 1	091318-DBRI-F-D-B	-58	-0.0028	-245.1		PPB	15 Aug 2018 12:24:43	U	SMPL
J4462-06 - 1	091318-DBRI-F-D-S	-41	0.0022	-245.1		PPB	15 Aug 2018 12:26:49	U	SMPL
J4462-07 - 1	091318-DBRI-F-U-B	-46	0.0007	-245.1		PPB	15 Aug 2018 12:28:57	U	SMPL
J4462-08 - 1	091318-DBRI-F-U-S	-31	0.0051	-245.1		PPB	15 Aug 2018 12:31:04	U	SMPL
J4458-01LX5 - 1		96	0.0422	-245.1		PPB	15 Aug 2018 12:33:12	U	SMPL
J4458-01A - 1		16502	4.8401	-245.1		PPB	15 Aug 2018 12:35:19	U	SMPL
CCV90 - 1	CCV90	16707	4.9000	-245.1		PPB	15 Aug 2018 12:37:27	U	SMPL
CCB90 - 1	CCB90	-156	-0.0315	-245.1		PPB	15 Aug 2018 12:39:37	U	SMPL