



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB97300

Ambient Barometric Pressure (mmHg): 755

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	27.0	80.6	80.9	08/15/2018	09:40
2	J4474-02		1	7	>100.0	>212.0	>212.0	08/15/2018	10:15
3	J4474-02DUP		1	7	>100.0	>212.0	>212.0	08/15/2018	10:46

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

LBq7300

WORKLIST(Hardcopy Internal Chain)

WorkList Name : fp81518

WorkList ID : 115653

Date : 8/15/2018 9:05:43 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/17/2018	Water	J4474-02 C	Flash Point	Cool 4 deg C	PSEG01	K63	FRAC-TANK-SV28570L	08/14/2018	1010A

Date/Time 08/15/18 9:10 AM
Received by: KS
Relinquished by: JP

Date/Time 08/15/18 10:50 AM
Received by: JP
Relinquished by: KS