



Analytical Summary Report

Analysis Method: SM 4500-PH B

Analyst By : sweta

Parameter: pH

Supervisor Review By : apatel

Run Number: LB97304

Slope : 99.0

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.9	4.01	08/15/2018	12:35
2	CAL2	1	Water			21.8	7.00	08/15/2018	12:37
3	CAL3	1	Water			22.0	10.04	08/15/2018	12:39
4	ICV	1	Water			21.8	7.00	08/15/2018	12:41
5	CCV1	1	Water			21.9	2.00	08/15/2018	12:43
6	J4489-01	1	Water			16.6	6.61	08/15/2018	12:45
7	J4489-01DUP	1	Water			16.7	6.62	08/15/2018	12:47
8	CCV2	1	Water			21.8	12.00	08/15/2018	12:49

WORKLIST(Hardcopy Internal Chain)

LB 97304

WorkList Name : pH081518

WorkList ID : 115669

Date : 8/15/2018 12:23:47 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/14/2018	Water	J4489-01	pH	Cool 4 deg C	PARS02	P53	SES-MS-PH4-COA-3QTR	08/14/2018	SM 4500-PH B

Date/Time 08/15/18 12:25
 Received by: SP
 Relinquished by: SP

Date/Time 08/15/18 1:05
 Received by: SP
 Relinquished by: SP