


7471B

LB97316
INSTRUMENT ID: CV7

Linear

 μ Abs.:

46382



A= 0.0000e+000

B= 2.1697e-004

C= 1.3116e-002

Rho= 0.9998795

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.023	0.023	47	0.000	47				
0.2	0.200	0.197	-0.003	848	0.0 %	848				
2.5	2.500	2.550	0.050	11692	0.0 %	11692				
5.0	5.000	4.923	-0.077	22630	0.0 %	22630				
7.5	7.500	7.430	-0.070	34185	0.0 %	34185				
10.0	10.000	10.076	0.076	46382	0.0 %	46382				

LB97316 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 15 Aug 2018 15:34:58

Sample ID	Extended ID	μ Abs.	Conc.	Std Cond	Method	Units	Date	Type	Type
0.0 - 1	50	47	-	0.00007471B	PPB		15 Aug 2018 15:38:08	S	Std
0.2 - 1	50.2	848	-	0.20007471B	PPB		15 Aug 2018 15:40:25	S	Std
2.5 - 1	52.5	11692	-	2.50007471B	PPB		15 Aug 2018 15:42:43	S	Std
5.0 - 1	55	22630	-	5.00007471B	PPB		15 Aug 2018 15:45:00	S	Std
7.5 - 1	57.5	34185	-	7.50007471B	PPB		15 Aug 2018 15:49:51	S	Std
10.0 - 1	57.5	46382	-	10.00007471B	PPB		15 Aug 2018 15:54:48	S	Std
ICV10 - 1	ICV10	18420	4.0096	-7471B	PPB		15 Aug 2018 15:57:28	U	SMPL
ICB10 - 1	ICB10	-126	-0.0142	-7471B	PPB		15 Aug 2018 15:59:44	U	SMPL
CCV93 - 1	CCV93	22940	4.9903	-7471B	PPB		15 Aug 2018 16:02:01	U	SMPL
CCB93 - 1	CCB93	-120	-0.0129	-7471B	PPB		15 Aug 2018 16:04:17	U	SMPL
CRA - 1	CRA	821	0.1912	-7471B	PPB		15 Aug 2018 16:06:33	U	SMPL
HighStd - 1	HighStd	46676	10.1403	-7471B	PPB		15 Aug 2018 16:08:49	U	SMPL
ChkStd - 1	ChkStd	31110	6.7630	-7471B	PPB		15 Aug 2018 16:11:05	U	SMPL
PB112063BL - 1	PBS	-218	-0.0342	-7471B	PPB		15 Aug 2018 16:13:24	U	SMPL
PB112063BS - 1	LCSS	16064	3.4985	-7471B	PPB		15 Aug 2018 16:15:42	U	SMPL
J3822-01 - 1	NFB-BF-SS001-01	144	0.0444	-7471B	PPB		15 Aug 2018 16:17:58	U	SMPL
J3822-02 - 1	NFB-BF-SS002-01	249	0.0671	-7471B	PPB		15 Aug 2018 16:20:16	U	SMPL
J3822-03 - 1	NFB-BF-SS003-01	233	0.0637	-7471B	PPB		15 Aug 2018 16:22:33	U	SMPL
J3822-04 - 1	NFB-BF-SS004-01	235	0.0641	-7471B	PPB		15 Aug 2018 16:24:51	U	SMPL
J3822-05 - 1	NFB-BF-SS005-01	500	0.1216	-7471B	PPB		15 Aug 2018 16:27:07	U	SMPL
CCV94 - 1	CCV94	22316	4.8550	-7471B	PPB		15 Aug 2018 16:29:22	U	SMPL
CCB94 - 1	CCB94	-134	-0.0160	-7471B	PPB		15 Aug 2018 16:31:38	U	SMPL
J4471-01 - 1	TP-1	3979	0.8764	-7471B	PPB		15 Aug 2018 16:33:55	U	SMPL
J4471-05 - 1	TP-2	762	0.1784	-7471B	PPB		15 Aug 2018 16:36:10	U	SMPL
J4471-09 - 1	TP-3	630	0.1498	-7471B	PPB		15 Aug 2018 16:38:27	U	SMPL
J4471-13 - 1	TP-4	585	0.1400	-7471B	PPB		15 Aug 2018 16:40:43	U	SMPL
J4471-17 - 1	TP-5	555	0.1335	-7471B	PPB		15 Aug 2018 16:42:59	U	SMPL
J4471-21 - 1	TP-6	812	0.1893	-7471B	PPB		15 Aug 2018 16:45:17	U	SMPL
J4471-25 - 1	BED-ROCK	2226	0.4961	-7471B	PPB		15 Aug 2018 16:47:34	U	SMPL
J4477-01 - 1	FK-SF-03-004-A	3043	0.6733	-7471B	PPB		15 Aug 2018 16:49:51	U	SMPL
J4477-03 - 1	FK-SF-03-004-B	3718	0.8198	-7471B	PPB		15 Aug 2018 16:52:09	U	SMPL
J4478-01 - 1	RICH-REC-ROAD-STONE-5	3300	0.7291	-7471B	PPB		15 Aug 2018 16:54:25	U	SMPL
CCV95 - 1	CCV95	22243	4.8391	-7471B	PPB		15 Aug 2018 16:56:41	U	SMPL
CCB95 - 1	CCB95	-121	-0.0131	-7471B	PPB		15 Aug 2018 16:58:58	U	SMPL
J4478-01DUP - 1	RICH-REC-ROAD-STONE-5DUP	3673	0.8100	-7471B	PPB		15 Aug 2018 17:01:14	U	SMPL
J4478-01MS - 1	RICH-REC-ROAD-STONE-5MS	20964	4.5616	-7471B	PPB		15 Aug 2018 17:03:30	U	SMPL
J4478-01MSD - 1	RICH-REC-ROAD-STONE-5MSD	20527	4.4668	-7471B	PPB		15 Aug 2018 17:05:46	U	SMPL
J4478-01LX5 - 1		540	0.1303	-7471B	PPB		15 Aug 2018 17:08:03	U	SMPL
J4478-01A - 1		21311	4.6369	-7471B	PPB		15 Aug 2018 17:10:21	U	SMPL
CCV96 - 1	CCV96	22517	4.8986	-7471B	PPB		15 Aug 2018 17:12:38	U	SMPL
CCB96 - 1	CCB96	-165	-0.0227	-7471B	PPB		15 Aug 2018 17:14:55	U	SMPL