



Analytical Summary Report

Analysis Method: SM 4500-PH B

Analyst By : sweta

Parameter: pH

Supervisor Review By : apatel

Run Number: LB97320

Slope : 98.6

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.9	4.01	08/15/2018	16:30
2	CAL2	1	Water			22.0	7.01	08/15/2018	16:32
3	CAL3	1	Water			21.8	10.04	08/15/2018	16:34
4	ICV	1	Water			21.9	7.00	08/15/2018	16:36
5	CCV1	1	Water			21.7	2.00	08/15/2018	16:38
6	J4505-01	1	Water			22.5	5.43	08/15/2018	16:40
7	J4505-01DUP	1	Water			22.6	5.44	08/15/2018	16:42
8	CCV2	1	Water			21.9	12.00	08/15/2018	16:48

2897320

WORKLIST(Hardcopy Internal Chain)

WorkList Name : pH081518-2

WorkList ID : 115685

Date : 8/15/2018 4:06:14 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/10/2018	Water	J4505-01B	pH	Cool 4 deg C	STER06	P33	POOL-WATER	08/10/2018	SM 4500-PH B

Date/Time 08/15/18 4:25
 Received by: SP
 Relinquished by: JP

Date/Time 08/15/18 4:50
 Received by: JP
 Relinquished by: SP