



Analytical Summary Report

Analysis Method: 9040C

Analyst By : sweta

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB97322

Slope : 98.6

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.9	4.01	08/15/2018	16:30
2	CAL2	1	Water			22.0	7.01	08/15/2018	16:32
3	CAL3	1	Water			21.8	10.04	08/15/2018	16:34
4	ICV	1	Water			21.9	7.00	08/15/2018	16:36
5	CCV1	1	Water			21.7	2.00	08/15/2018	16:38
6	J4502-04	1	Water			26.8	6.77	08/15/2018	16:44
7	J4502-04DUP	1	Water			26.9	6.78	08/15/2018	16:46
8	CCV2	1	Water			21.9	12.00	08/15/2018	16:48

LB 97322

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIVITY081518-1

WorkList ID : 115680

Date : 8/15/2018 3:36:31 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/15/2018	Water	J4502-04	Corrosivity	Cool 4 deg C	PSEG01	K63	BROWN-FRAC-TANK	08/15/2018	9040C

Date/Time 08/15/18 4:25
Received by: SP
Relinquished by: JP

Date/Time 08/15/18 4:50
Received by: JP
Relinquished by: SP