

**TOTAL SUSPENDED SOLIDS - SM2540D**SUPERVISOR: apatelANALYST: AparnaDate: 08/16/2018TEMP1 IN: 103 08/16/2018 07:10 TEMP OUT: 104 08/16/2018 08:15TEMP2 IN: 104 08/16/2018 09:00 TEMP OUT: 104 08/16/2018 10:05TEMP3 IN: 105 08/16/2018 11:30 TEMP OUT: 105 08/16/2018 13:05TEMP4 IN: 105 08/16/2018 13:40 TEMP OUT: 105 08/16/2018 15:20Run Number: LB97330BalanceID: WC SC-3OvenID: WC Oven-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L
1	LB97330BL	LB97330BL	1.4020	1.4020	100	1.4020	1.4020	1.4020	0.0000	0
2	LB97330BS	LB97330BS	1.4056	1.4056	100	1.4560	1.4560	1.4560	0.0504	504
3	J4473-02	KY038PS027-180813	1.3939	1.3939	900	1.3943	1.3943	1.3943	0.0004	0.4
4	J4489-01	SES-MS-PH4-COA-3QTR	1.4016	1.4016	800	1.4020	1.4020	1.4020	0.0004	0.5
5	J4502-03	BROWN-FRAC-TANK	1.3957	1.3957	25	1.4130	1.4130	1.4130	0.0173	692
6	J4502-03DUP	BROWN-FRAC-TANKDUP	1.3891	1.3891	25	1.4069	1.4069	1.4069	0.0178	712
7	J4504-01	081418-DBRI-F-D-B	1.4133	1.4133	300	1.4195	1.4195	1.4195	0.0062	20.7
8	J4504-02	081418-DBRI-F-D-S	1.3976	1.3976	400	1.4037	1.4037	1.4037	0.0061	15.3
9	J4504-03	081418-DBRI-F-U-B	1.3909	1.3909	400	1.3989	1.3989	1.3989	0.0080	20
10	J4504-04	081418-DBRI-F-U-S	1.4035	1.4035	400	1.4091	1.4091	1.4091	0.0056	14
11	J4504-05	081418-TIPC-F-D-B	1.3912	1.3912	300	1.4024	1.4024	1.4024	0.0112	37.3
12	J4504-06	081418-TIPC-F-D-S	1.4065	1.4065	300	1.4131	1.4131	1.4131	0.0066	22
13	J4504-07	081418-TIPC-F-U-B	1.4117	1.4117	300	1.4224	1.4224	1.4224	0.0107	35.7
14	J4504-08	081418-TIPC-F-U-S	1.4052	1.4052	200	1.4124	1.4124	1.4124	0.0072	36
15	J4504-08DUP	081418-TIPC-F-U-SDUP	1.4114	1.4114	200	1.4188	1.4188	1.4188	0.0074	37



TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: apatel

ANALYST: Aparna

Date: 08/16/2018

TEMP1 IN: 103 08/16/2018 07:10 TEMP OUT: 104 08/16/2018 08:15

TEMP2 IN: 104 08/16/2018 09:00 TEMP OUT: 104 08/16/2018 10:05

TEMP3 IN: 105 08/16/2018 11:30 TEMP OUT: 105 08/16/2018 13:05

TEMP4 IN: 105 08/16/2018 13:40 TEMP OUT: 105 08/16/2018 15:20

Run Number: LB97330

BalanceID: WC SC-3

OvenID: WC Oven-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L
16	J4506-02	EFF-WASTE-WATER	1.4027	1.4027	300	1.4058	1.4058	1.4058	0.0031	10.3

A = Sample Volume (ml)

B = Empty Dish Weight (g)

C = Final Empty Dish + Sample weight after 1.5 hr drying @105°C(g)

D = Weight (g)

Weight (g) = C - B

Result mg/L = $\frac{D}{A} \times 1000 \times 1000$

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TSS081618

WorkList ID : 115691

Date : 8/16/2018 7:49:23 AM

1397330

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/20/2018	Water	J4473-02 D	TSS	Cool 4 deg C	SHAW22	P31	KY038PS027-180813	08/13/2018	SM2540 D
08/21/2018	Water	J4489-01 G	TSS	Cool 4 deg C	PARS02	P53	SES-MS-PH4-COA-3QTR	08/14/2018	SM2540 D
08/20/2018	Water	J4502-03 G	TSS	Cool 4 deg C	PSEG01	K63	BROWN-FRAC-TANK	08/15/2018	SM2540 D
08/21/2018	Water	J4504-01 C	TSS	Cool 4 deg C	HDRI03	P22	081418-DBRI-F-D-B	08/14/2018	SM2540 D
08/21/2018	Water	J4504-02 C	TSS	Cool 4 deg C	HDRI03	P22	081418-DBRI-F-D-S	08/14/2018	SM2540 D
08/21/2018	Water	J4504-03 C	TSS	Cool 4 deg C	HDRI03	P22	081418-DBRI-F-U-B	08/14/2018	SM2540 D
08/21/2018	Water	J4504-04 C	TSS	Cool 4 deg C	HDRI03	P22	081418-DBRI-F-U-S	08/14/2018	SM2540 D
08/17/2018	Water	J4504-05 C	TSS	Cool 4 deg C	HDRI03	P22	081418-TIPC-F-D-B	08/14/2018	SM2540 D
08/17/2018	Water	J4504-06 C	TSS	Cool 4 deg C	HDRI03	P22	081418-TIPC-F-D-S	08/14/2018	SM2540 D
08/17/2018	Water	J4504-07 C	TSS	Cool 4 deg C	HDRI03	P22	081418-TIPC-F-U-B	08/14/2018	SM2540 D
08/17/2018	Water	J4504-08 C	TSS	Cool 4 deg C	HDRI03	P22	081418-TIPC-F-U-S	08/14/2018	SM2540 D
08/22/2018	Water	J4506-02 B	TSS	Cool 4 deg C	ARDM01	P63	EFF-WASTE-WATER	08/15/2018	SM2540 D

FP-9823888

Date/Time 08/16/18, 08:00
 Received by: AS
 Relinquished by: AS

Date/Time 08/16/18, 12:00
 Received by: AS
 Relinquished by: AS