

TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: apatel

ANALYST: Aparna

Date: 08/06/2018

TEMP1 IN: 103 08/06/2018 07:05 TEMP OUT: 104 08/06/2018 08:15

TEMP2 IN: 104 08/06/2018 09:00 TEMP OUT: 104 08/06/2018 10:05

TEMP3 IN: 105 08/06/2018 12:00 TEMP OUT: 105 08/06/2018 13:35

TEMP4 IN: 105 08/06/2018 14:11 TEMP OUT: 105 08/06/2018 15:40

Run Number: LB97334

BalanceID: WC SC-3

OvenID: WC Oven-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L
1	LB97334BL	LB97334BL	1.4021	1.4021	100	1.4021	1.4021	1.4021	0.0000	0
2	LB97334BS	LB97334BS	1.4012	1.4012	100	1.4515	1.4515	1.4515	0.0503	503
3	J4350-01	080218-DBRI-E-D-B	1.4091	1.4091	300	1.4291	1.4291	1.4291	0.0200	66.7
4	J4350-02	080218-DBRI-E-D-S	1.3998	1.3998	300	1.4150	1.4150	1.4150	0.0152	50.7
5	J4350-03	080218-DBRI-E-U-B	1.4021	1.4021	200	1.4183	1.4183	1.4183	0.0162	81
6	J4350-04	080218-DBRI-E-U-S	1.3910	1.3910	300	1.4056	1.4056	1.4056	0.0146	48.7
7	J4350-05	080218-TIPC-E-D-M	1.4034	1.4034	300	1.4215	1.4215	1.4215	0.0181	60.3
8	J4350-06	080218-TIPC-E-U-M	1.4011	1.4011	200	1.4120	1.4120	1.4120	0.0109	54.5
9	J4350-06DUP	080218-TIPC-E-U-MDUP	1.3948	1.3948	200	1.4059	1.4059	1.4059	0.0111	55.5
10	J4352-01	GROUNDWATER	1.3995	1.3995	100	1.4323	1.4323	1.4323	0.0328	328

A = Sample Volume (ml)

B = Empty Dish Weight (g)

C = Final Empty Dish + Sample weight after 1.5 hr drying @105°C(g)

D = Weight (g)

Weight (g) = C - B

Result mg/L = $\frac{D}{A} \times 1000 \times 1000$

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TSS080618-3

WorkList ID : 115714

Date : 8/06/2018 10:01:41 AM

2B9734

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/09/2018	Water	J4350-01 C	TSS	Cool 4 deg C	HDRI03	Q32	080218-DBRI-E-D-B	08/02/2018	SM2540 D
08/09/2018	Water	J4350-02 C	TSS	Cool 4 deg C	HDRI03	Q32	080218-DBRI-E-D-S	08/02/2018	SM2540 D
08/09/2018	Water	J4350-03 C	TSS	Cool 4 deg C	HDRI03	Q32	080218-DBRI-E-U-B	08/02/2018	SM2540 D
08/09/2018	Water	J4350-04 C	TSS	Cool 4 deg C	HDRI03	Q32	080218-DBRI-E-U-S	08/02/2018	SM2540 D
08/07/2018	Water	J4350-05 C	TSS	Cool 4 deg C	HDRI03	Q32	080218-TIPC-E-D-M	08/02/2018	SM2540 D
08/07/2018	Water	J4350-06 C	TSS	Cool 4 deg C	HDRI03	Q32	080218-TIPC-E-U-M	08/02/2018	SM2540 D
08/10/2018	Water	J4352-01 A	TSS	Cool 4 deg C	SEAC01	Q32	GROUNDWATER	08/03/2018	SM2540 D

FP-9823888

Date/Time 08/06/18, 08:00
 Received by: AS
 Relinquished by: AS

Date/Time 08/06/18, 12:00
 Received by: AS
 Relinquished by: AS