



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB97337

Ambient Barometric Pressure (mmHg): 760

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.0	80.6	80.6	08/16/2018	09:40
2	J4502-04		1	8	>100.0	>212.0	>212.0	08/16/2018	10:15
3	J4502-04DUP		1	8	>100.0	>212.0	>212.0	08/16/2018	10:40

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

LB 97337

WORKLIST(Hardcopy Internal Chain)

WorkList Name : FP81618

WorkList ID : 115712

Date : 8/16/2018 9:07:06 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/20/2018	Water	J4502-04 C	Flash Point	Cool 4 deg C	PSEG01	K63	BROWN-FRAC-TANK	08/15/2018	1010A

Date/Time 08/16/18 9:10 AM

Received by: KS

Relinquished by: RJ

Date/Time 08/16/18 10:45 AM

Received by: RJ

Relinquished by: KS