

LB97338

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

8/16/2018 10:44

Test: Phenolics-*mg/d*

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	1.045	0.0	0.112	
ICB1	0.007	0.0	0.004	
CCV1	0.979	0.0	0.105	
CCB1	0.005	0.0	0.004	
PB112047BL	0.006	0.0	0.004	
PB112047BS	1.051	0.0	0.113	
J4458-01	0.037	0.0	0.007	
J4458-01DUP	0.035	0.0	0.007	
J4458-01MS	1.007	0.0	0.108	
J4458-01MSD	1.024	0.0	0.110	
J4458-02	0.009	0.0	0.004	
J4458-03	0.013	0.0	0.005	
CCV2	0.972	0.0	0.105	
CCB2	0.005	0.0	0.004	

N	14
Mean	0.443
SD	0.5132
CV%	115.98

AS

Aquakem v. 7.2AQ1

Results from time period:

Thu Aug 16 09:20:06 2018

Thu Aug 16 09:52:15 2018

Sample Id	Sam/Ctr/c	Test short nam	Test type	Result	Result unit	Result date and time	Stat
0.05PPM	A	Phenolics-	P	0.0541	mg/l	8/16/2018 9:20:06	
0.1PPM	A	Phenolics-	P	0.1012	mg/l	8/16/2018 9:20:07	
0.25PPM	A	Phenolics-	P	0.2499	mg/l	8/16/2018 9:20:08	
0.50PPM	A	Phenolics-	P	0.497	mg/l	8/16/2018 9:20:09	
1.0PPM	A	Phenolics-	P	0.9945	mg/l	8/16/2018 9:20:10	
2.0PPM	A	Phenolics-	P	2.0034	mg/l	8/16/2018 9:20:11	
ICV1	S	Phenolics-	P	1.0447	mg/l	8/16/2018 9:47:17	
ICB1	S	Phenolics-	P	0.0072	mg/l	8/16/2018 9:47:18	
CCV1	S	Phenolics-	P	0.9788	mg/l	8/16/2018 9:47:19	
CCB1	S	Phenolics-	P	0.0054	mg/l	8/16/2018 9:47:20	
PB112047BL	S	Phenolics-	P	0.0065	mg/l	8/16/2018 9:47:21	
PB112047BS	S	Phenolics-	P	1.0505	mg/l	8/16/2018 9:47:22	
J4458-01	S	Phenolics-	P	0.0372	mg/l	8/16/2018 9:47:23	
J4458-01DUP	S	Phenolics-	P	0.0347	mg/l	8/16/2018 9:47:24	
J4458-01MS	S	Phenolics-	P	1.0071	mg/l	8/16/2018 9:47:25	
J4458-01MSD	S	Phenolics-	P	1.0245	mg/l	8/16/2018 9:47:26	
J4458-02	S	Phenolics-	P	0.0087	mg/l	8/16/2018 9:47:27	
J4458-03	S	Phenolics-	P	0.0129	mg/l	8/16/2018 9:47:28	
CCV2	S	Phenolics-	P	0.9725	mg/l	8/16/2018 9:52:12	
CCB2	S	Phenolics-	P	0.0047	mg/l	8/16/2018 9:52:13	

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

8/16/2018 9:21

Test Phenolics-


Accepted 8/16/2018 9:21

Factor 9.587

Bias 0.003

Coeff. of det. 0.999975

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	Phenol-2	0.009	0.0541	0.0500	
2	Phenol-2	0.014	0.1012	0.1000	
3	Phenol-2	0.029	0.2499	0.2500	
4	Phenol-2	0.055	0.4970	0.5000	
5	Phenol-2	0.107	0.9945	1.0000	
6	Phenol-2	0.212	2.0034	2.0000	