

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

8/16/2018 11:10

Test: Phenolics-mg/l

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	1.046	0.0	0.112	
ICB1	0.002	0.0	0.004	
CCV1	0.980	0.0	0.106	
CCB1	0.002	0.0	0.004	
PB112092BL	0.002	0.0	0.004	
PB112092BS	1.063	0.0	0.114	
J4489-01	0.005	0.0	0.004	
J4489-01DUP	0.005	0.0	0.004	
J4489-01MS	0.898	0.0	0.097	
J4489-01MSD	0.888	0.0	0.096	
CCV2	0.985	0.0	0.106	
CCB2	0.001	0.0	0.003	

N	12
Mean	0.490
SD	0.5110
CV%	104.34

AS

Aquakem v. 7.2AQ1

Results from time period:

Thu Aug 16 11:08:15 2018

Thu Aug 16 11:08:26 2018

Sample Id	Sam/Ctr/c	Test short nar	Test type	Result	Result unit	Result date and time	Stat
0.05PPM	A	Phenolics-	P	0.0541	mg/l	8/16/2018 9:20:06	
0.1PPM	A	Phenolics-	P	0.1012	mg/l	8/16/2018 9:20:07	
0.25PPM	A	Phenolics-	P	0.2499	mg/l	8/16/2018 9:20:08	
0.50PPM	A	Phenolics-	P	0.497	mg/l	8/16/2018 9:20:09	
1.0PPM	A	Phenolics-	P	0.9945	mg/l	8/16/2018 9:20:10	
2.0PPM	A	Phenolics-	P	2.0034	mg/l	8/16/2018 9:20:11	
ICV1	S	Phenolics-	P	1.0463	mg/l	8/16/2018 11:08:15	
ICB1	S	Phenolics-	P	0.0021	mg/l	8/16/2018 11:08:16	
CCV1	S	Phenolics-	P	0.9799	mg/l	8/16/2018 11:08:17	
CCB1	S	Phenolics-	P	0.0021	mg/l	8/16/2018 11:08:18	
PB112092BL	S	Phenolics-	P	0.0018	mg/l	8/16/2018 11:08:19	
PB112092BS	S	Phenolics-	P	1.0635	mg/l	8/16/2018 11:08:20	
J4489-01	S	Phenolics-	P	0.0053	mg/l	8/16/2018 11:08:21	
J4489-01DUP	S	Phenolics-	P	0.0048	mg/l	8/16/2018 11:08:22	
J4489-01MS	S	Phenolics-	P	0.8978	mg/l	8/16/2018 11:08:23	
J4489-01MSD	S	Phenolics-	P	0.8878	mg/l	8/16/2018 11:08:24	
CCV2	S	Phenolics-	P	0.9854	mg/l	8/16/2018 11:08:25	
CCB2	S	Phenolics-	P	0.0005	mg/l	8/16/2018 11:08:26	

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

8/16/2018 9:21


Test Phenolics-

Accepted 8/16/2018 9:21

Factor 9.587
Bias 0.003

Coeff. of det. 0.999975

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	Phenol-2	0.009	0.0541	0.0500	
2	Phenol-2	0.014	0.1012	0.1000	
3	Phenol-2	0.029	0.2499	0.2500	
4	Phenol-2	0.055	0.4970	0.5000	
5	Phenol-2	0.107	0.9945	1.0000	
6	Phenol-2	0.212	2.0034	2.0000	