

LB97340

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

8/15/2018 16:09

Test: ^{Reactive} ~~Total~~ CN ^{µg/d}

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	97.135	0.0	0.093	
ICB1	-0.240	0.0	0.004	
CCV1	244.850	0.0	0.228	
CCB1	-0.261	0.0	0.004	
PB112074BL	-0.919	0.0	0.004	
J4471-04	-0.669	0.0	0.004	
J4471-04DUP	-0.780	0.0	0.004	
J4471-08	-0.547	0.0	0.004	
J4471-12	-0.759	0.0	0.004	
J4471-16	-0.872	0.0	0.004	
J4471-20	-0.746	0.0	0.004	
J4471-24	-0.715	0.0	0.004	
J4471-27	-0.977	0.0	0.004	
PB112075BL	-0.306	0.0	0.004	
CCV2	244.302	0.0	0.227	
CCB2	-0.426	0.0	0.004	
J4474-02	-0.857	0.0	0.004	
J4474-02DUP	-0.402	0.0	0.004	
CCV3	247.914	0.0	0.231	
CCB3	-0.149	0.0	0.004	

N	20
Mean	41.229
SD	90.7693
CV%	220.16

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AD

8/15/2018 14:47


Test Total CN

Accepted 8/15/2018 14:47

Factor 1096
Bias 0.004

Coeff. of det. 0.999917

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.004	-0.6851	0.0000	
2	5.0PPBCN	0.008	4.1873	5.0000	
3	10PPBCN	0.013	9.2814	10.0000	
4	50PPBCN	0.050	50.0697	50.0000	
5	100PPBCN	0.096	100.5089	100.0000	
6	250PPBCN	0.236	253.4482	250.0000	
7	500PPBCN	0.459	498.1896	500.0000	

Aquakem v. 7.2AQ1

Results from time period:

Wed Aug 15 16:01:33 2018

Wed Aug 15 16:08:20 2018

AS

Sample Id	Sam/Ctr/c/	Test short nam	Test type	Result	Result unit	Result date and time	Stat
0.0PPBCN	A	Reactive CN	P	-0.6851	µg/l	8/15/2018 14:46:45	
5.0PPBCN	A	Reactive CN	P	4.1873	µg/l	8/15/2018 14:46:46	
10PPBCN	A	Reactive CN	P	9.2814	µg/l	8/15/2018 14:46:47	
50PPBCN	A	Reactive CN	P	50.0697	µg/l	8/15/2018 14:46:48	
100PPBCN	A	Reactive CN	P	100.5089	µg/l	8/15/2018 14:46:49	
250PPBCN	A	Reactive CN	P	253.4482	µg/l	8/15/2018 14:46:50	
500PPBCN	A	Reactive CN	P	498.1896	µg/l	8/15/2018 14:46:51	
ICV1	S	Reactive CN	P	97.1347	µg/l	8/15/2018 16:01:33	
ICB1	S	Reactive CN	P	-0.2398	µg/l	8/15/2018 16:01:34	
CCV1	S	Reactive CN	P	244.8502	µg/l	8/15/2018 16:01:35	
CCB1	S	Reactive CN	P	-0.261	µg/l	8/15/2018 16:01:36	
PB112074BL	S	Reactive CN	P	-0.9187	µg/l	8/15/2018 16:01:37	
J4471-04	S	Reactive CN	P	-0.6687	µg/l	8/15/2018 16:01:38	
J4471-04DUP	S	Reactive CN	P	-0.7804	µg/l	8/15/2018 16:01:39	
J4471-08	S	Reactive CN	P	-0.547	µg/l	8/15/2018 16:01:40	
J4471-12	S	Reactive CN	P	-0.7594	µg/l	8/15/2018 16:01:41	
J4471-16	S	Reactive CN	P	-0.8722	µg/l	8/15/2018 16:01:42	
J4471-20	S	Reactive CN	P	-0.7462	µg/l	8/15/2018 16:01:43	
J4471-24	S	Reactive CN	P	-0.7148	µg/l	8/15/2018 16:08:12	
J4471-27	S	Reactive CN	P	-0.9767	µg/l	8/15/2018 16:08:13	
PB112075BL	S	Reactive CN	P	-0.3064	µg/l	8/15/2018 16:08:14	
CCV2	S	Reactive CN	P	244.302	µg/l	8/15/2018 16:08:15	
CCB2	S	Reactive CN	P	-0.4256	µg/l	8/15/2018 16:08:16	
J4474-02	S	Reactive CN	P	-0.8574	µg/l	8/15/2018 16:08:17	
J4474-02DUP	S	Reactive CN	P	-0.4023	µg/l	8/15/2018 16:08:18	
CCV3	S	Reactive CN	P	247.9136	µg/l	8/15/2018 16:08:19	
CCB3	S	Reactive CN	P	-0.1494	µg/l	8/15/2018 16:08:20	