

LB 97343


7471B

INSTRUMENT ID: CV 1

Linear

 μ Abs.:

44209



A= 0.0000e+000

B= 2.2819e-004

C= -3.8268e-002

Rho= 0.9999022

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.021	-0.021	74	0.000	74				
0.2	0.200	0.194	-0.006	1019	0.0 %	1019				
2.5	2.500	2.568	0.068	11423	0.0 %	11423				
5.0	5.000	5.000	0.000	22081	0.0 %	22081				
7.5	7.500	7.409	-0.091	32635	0.0 %	32635				
10.0	10.000	10.050	0.050	44209	0.0 %	44209				

LB97343 INSTRUMENT ID : CV1

Method: 7471B		Operator: Admin		Date of Analysis: 16 Aug 2018 13:07:46					
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	74	-	0.0000	7471B	PPB	16 Aug 2018 13:47:57	S	Std
0.2 - 1	50.2	1019	-	0.2000	7471B	PPB	16 Aug 2018 13:50:15	S	Std
2.5 - 1	52.5	11423	-	2.5000	7471B	PPB	16 Aug 2018 13:52:32	S	Std
5.0 - 1	55	22081	-	5.0000	7471B	PPB	16 Aug 2018 13:54:50	S	Std
7.5 - 1	57.5	32635	-	7.5000	7471B	PPB	16 Aug 2018 13:57:07	S	Std
10.0 - 1	510	44209	-	10.0000	7471B	PPB	16 Aug 2018 13:59:24	S	Std
ICV12 - 1	ICV12	18205	4.1159	-	7471B	PPB	16 Aug 2018 14:02:04	U	SMPL
ICB12 - 1	ICB12	127	-0.0093	-	7471B	PPB	16 Aug 2018 14:04:21	U	SMPL
CCV03 - 1	CCV03	22568	5.1115	-	7471B	PPB	16 Aug 2018 14:06:38	U	SMPL
CCB03 - 1	CCB03	51	-0.0266	-	7471B	PPB	16 Aug 2018 14:08:54	U	SMPL
CRA - 1	CRA	1070	0.2059	-	7471B	PPB	16 Aug 2018 14:11:11	U	SMPL
HighStd - 1	HighStd	44796	10.1837	-	7471B	PPB	16 Aug 2018 14:13:27	U	SMPL
ChkStd - 1	ChkStd	31518	7.1538	-	7471B	PPB	16 Aug 2018 14:15:43	U	SMPL
PB112101BL - 1	PBS	28	-0.0319	-	7471B	PPB	16 Aug 2018 14:20:37	U	SMPL
PB112101BS - 1	LCSS	17043	3.8508	-	7471B	PPB	16 Aug 2018 14:22:53	U	SMPL
J4499-01 - 1	LSB-01(0-1)	36208	8.2240	-	7471B	PPB	16 Aug 2018 14:25:11	U	SMPL
J4499-02 - 1	LSB-02(0-1)	23391	5.2993	-	7471B	PPB	16 Aug 2018 14:27:29	U	SMPL
J4499-03 - 1	LSB-03(0-1)	11426	2.5690	-	7471B	PPB	16 Aug 2018 14:29:49	U	SMPL
J4499-04 - 1	LSB-04(0-1)	20504	4.6405	-	7471B	PPB	16 Aug 2018 14:32:09	U	SMPL
J4499-05 - 1	LSB-05(0-1)	44131	10.0319	-	7471B	PPB	16 Aug 2018 14:34:25	U	SMPL
CCV04 - 1	CCV04	21799	4.9360	-	7471B	PPB	16 Aug 2018 14:36:42	U	SMPL
CCB04 - 1	CCB04	-5	-0.0394	-	7471B	PPB	16 Aug 2018 14:38:59	U	SMPL
J4499-06 - 1	LSB-06(0-1)	6676	1.4851	-	7471B	PPB	16 Aug 2018 14:41:16	U	SMPL
J4499-07 - 1	LSB-07(0-1)	25389	5.7552	-	7471B	PPB	16 Aug 2018 14:43:32	U	SMPL
J4499-07DUP - 1	LSB-07(0-1)DUP	23698	5.3694	-	7471B	PPB	16 Aug 2018 14:45:49	U	SMPL
J4499-08 - 1	J4499-07MS	39934	9.0742	-	7471B	PPB	16 Aug 2018 14:48:07	U	SMPL
J4499-09 - 1	J4499-07MSD	42689	9.7029	-	7471B	PPB	16 Aug 2018 14:52:59	U	SMPL
J4499-10 - 1	LSB-08(0-1)	2685	0.5744	-	7471B	PPB	16 Aug 2018 14:55:15	U	SMPL
J4499-11 - 1	DUP(0-1)	45057	10.2432	-	7471B	PPB	16 Aug 2018 14:57:34	U	SMPL
J4501-02 - 1	RA-081318-FL-017	16819	3.7996	-	7471B	PPB	16 Aug 2018 14:59:52	U	SMPL
J4501-04 - 1	RA-081318-FL-019	7865	1.7564	-	7471B	PPB	16 Aug 2018 15:02:11	U	SMPL
J4501-05 - 1	RA-081318-FL-020	2568	0.5477	-	7471B	PPB	16 Aug 2018 15:04:27	U	SMPL
CCV05 - 1	CCV05	22432	5.0805	-	7471B	PPB	16 Aug 2018 15:06:43	U	SMPL
CCB05 - 1	CCB05	38	-0.0296	-	7471B	PPB	16 Aug 2018 15:08:59	U	SMPL
J4501-06 - 1	RA-081318-SSW-15-021	2847	0.6114	-	7471B	PPB	16 Aug 2018 15:11:16	U	SMPL
J4501-07 - 1	RA-081318-SSW-45-022	4255	0.9327	-	7471B	PPB	16 Aug 2018 15:13:32	U	SMPL
PB112102BL - 1	PBS	170	0.0005	-	7471B	PPB	16 Aug 2018 15:23:48	U	SMPL
PB112102BS - 1	LCSS	16332	3.6885	-	7471B	PPB	16 Aug 2018 15:26:04	U	SMPL
J4269-01 - 1	AU-06-081518-A	2532	0.5395	-	7471B	PPB	16 Aug 2018 15:33:17	U	SMPL
J4269-01DUP - 1	AU-06-081518-ADUP	2496	0.5313	-	7471B	PPB	16 Aug 2018 15:35:33	U	SMPL
J4269-01MS - 1	AU-06-081518-AMS	19270	4.3589	-	7471B	PPB	16 Aug 2018 15:37:51	U	SMPL
J4269-01MSD - 1	AU-06-081518-AMSD	19303	4.3665	-	7471B	PPB	16 Aug 2018 15:40:08	U	SMPL
J4269-03 - 1	AU-06-081518-B	2770	0.5938	-	7471B	PPB	16 Aug 2018 15:42:27	U	SMPL
J4282-05 - 1	SU-01-081518-A	1356	0.2712	-	7471B	PPB	16 Aug 2018 15:44:43	U	SMPL
CCV06 - 1	CCV06	22243	5.0373	-	7471B	PPB	16 Aug 2018 15:46:59	U	SMPL
CCB06 - 1	CCB06	139	-0.0065	-	7471B	PPB	16 Aug 2018 15:49:15	U	SMPL
J4282-07 - 1	SU-01-081518-B	1667	0.3421	-	7471B	PPB	16 Aug 2018 15:51:31	U	SMPL
J4483-02 - 1	TOD1-IB-0-4	4400	0.9658	-	7471B	PPB	16 Aug 2018 15:53:48	U	SMPL
J4486-01 - 1	VNJ202	109886	25.0365	-	7471B	PPB	16 Aug 2018 15:56:05	U	SMPL
J4496-01 - 1	CARLSTADT-TOP-SOIL	7692	1.7170	-	7471B	PPB	16 Aug 2018 15:58:21	U	SMPL
J4497-01 - 1	CARLSTADT-FRONT	7148	1.5928	-	7471B	PPB	16 Aug 2018 16:00:43	U	SMPL
J4498-01 - 1	CARLSTADT-BACK	27101	6.1459	-	7471B	PPB	16 Aug 2018 16:03:00	U	SMPL
J4500-01 - 1	TP-1	1405	0.2823	-	7471B	PPB	16 Aug 2018 16:05:17	U	SMPL
J4500-05 - 1	TP-2	1417	0.2851	-	7471B	PPB	16 Aug 2018 16:07:37	U	SMPL
J4500-09 - 1	TP-6	1272	0.2520	-	7471B	PPB	16 Aug 2018 16:09:54	U	SMPL
J4500-13 - 1	TP-7	2268	0.4793	-	7471B	PPB	16 Aug 2018 16:12:10	U	SMPL
CCV07 - 1	CCV07	22494	5.0946	-	7471B	PPB	16 Aug 2018 16:14:27	U	SMPL
CCB07 - 1	CCB07	122	-0.0104	-	7471B	PPB	16 Aug 2018 16:16:43	U	SMPL
J4500-17 - 1	TP-12	2924	0.6290	-	7471B	PPB	16 Aug 2018 16:19:01	U	SMPL
J4500-21 - 1	ROCK	953	0.1792	-	7471B	PPB	16 Aug 2018 16:21:17	U	SMPL
J4502-01 - 1	LGS-CONCRETE	665	0.1135	-	7471B	PPB	16 Aug 2018 16:23:33	U	SMPL
J4499-07LX5 - 1		4911	1.0824	-	7471B	PPB	16 Aug 2018 16:25:50	U	SMPL
J4499-07A - 1		42311	9.6166	-	7471B	PPB	16 Aug 2018 16:28:06	U	SMPL
J4269-01LX5 - 1		480	0.0713	-	7471B	PPB	16 Aug 2018 16:30:23	U	SMPL
J4269-01A - 1		20793	4.7065	-	7471B	PPB	16 Aug 2018 16:32:43	U	SMPL
J4499-05DLX2 - 1		22102	5.0052	-	7471B	PPB	16 Aug 2018 16:35:00	U	SMPL
J4499-11DLX2 - 1		22740	5.1507	-	7471B	PPB	16 Aug 2018 16:37:18	U	SMPL
J4486-01DLX5 - 1		22019	4.9862	-	7471B	PPB	16 Aug 2018 16:39:35	U	SMPL
CCV08 - 1	CCV08	22381	5.0688	-	7471B	PPB	16 Aug 2018 16:41:54	U	SMPL
CCB08 - 1	CCB08	74	-0.0214	-	7471B	PPB	16 Aug 2018 16:44:11	U	SMPL