



Analytical Summary Report

Analysis Method: 9040C

Analyst By : sweta

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB97355

Slope : 98.6

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.9	4.01	08/16/2018	16:10
2	CAL2	1	Water			21.7	7.01	08/16/2018	16:12
3	CAL3	1	Water			21.7	10.04	08/16/2018	16:14
4	ICV	1	Water			21.8	7.00	08/16/2018	16:16
5	CCV1	1	Water			22.0	2.00	08/16/2018	16:18
6	J4516-07	1	Water			24.4	6.81	08/16/2018	16:20
7	J4516-07DUP	1	Water			24.5	6.82	08/16/2018	16:22
8	CCV2	1	Water			21.8	12.00	08/16/2018	16:24

LB 97355

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIVITY081618-1

WorkList ID : 115731

Date : 8/16/2018 3:49:01 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/16/2018	Water	J4516-07	Corrosivity	Cool 4 deg C	PSEG01	K63	ND-BASELINE-WATER	08/16/2018	9040C

Date/Time 08/16/18 4:00
Received by: SP
Relinquished by: SP

Date/Time 08/16/18 4:30
Received by: SP
Relinquished by: SP