

245.1


LB 97357

INSTRUMENT ID: CV1

Linear

 μ Abs.:

33098



A= 0.0000e+000

B= 3.0225e-004

C= -1.6001e-002

Rho= 0.9998882

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.014	-0.014	5	0.000	5				
0.05	0.050	0.065	0.015	268	0.0 %	268				
0.2	0.200	0.225	0.025	799	0.0 %	799				
2.5	2.500	2.518	0.018	8383	0.0 %	8383				
5.0	5.000	4.885	-0.115	16216	0.0 %	16216				
7.5	7.500	7.583	0.083	25141	0.0 %	25141				
10.0	10.000	9.988	-0.012	33098	0.0 %	33098				

LB97357 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 17 Aug 2018 08:55:45

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	5	-	0.0000	245.1	PPB	17 Aug 2018 09:02:12	S	Std
0.05 - 1	50.05	268	-	0.0500	245.1	PPB	17 Aug 2018 09:06:47	S	Std
0.2 - 1	50.2	799	-	0.2000	245.1	PPB	17 Aug 2018 09:08:55	S	Std
2.5 - 1	52.5	8383	-	2.5000	245.1	PPB	17 Aug 2018 09:11:03	S	Std
5.0 - 1	55	16216	-	5.0000	245.1	PPB	17 Aug 2018 09:13:11	S	Std
7.5 - 1	57.5	25141	-	7.5000	245.1	PPB	17 Aug 2018 09:15:18	S	Std
10.0 - 1	510	33098	-	10.0000	245.1	PPB	17 Aug 2018 09:17:27	S	Std
ICV13 - 1	ICV13	13597	4.0937	-245.1		PPB	17 Aug 2018 09:20:10	U	SMPL
ICB13 - 1	ICB13	31	-0.0066	-245.1		PPB	17 Aug 2018 09:22:17	U	SMPL
CCV09 - 1	CCV09	16468	4.9615	-245.1		PPB	17 Aug 2018 09:24:25	U	SMPL
CCB09 - 1	CCB09	33	-0.0060	-245.1		PPB	17 Aug 2018 09:26:31	U	SMPL
CRA - 1	CRA	231	0.0538	-245.1		PPB	17 Aug 2018 09:28:40	U	SMPL
HighStd - 1	HighStd	33145	10.0022	-245.1		PPB	17 Aug 2018 09:30:46	U	SMPL
ChkStd - 1	ChkStd	23717	7.1525	-245.1		PPB	17 Aug 2018 09:32:53	U	SMPL
PB112111BL - 1	PBW	97	0.0133	-245.1		PPB	17 Aug 2018 09:37:35	U	SMPL
PB112111BS - 1	LCSW	13475	4.0569	-245.1		PPB	17 Aug 2018 09:39:42	U	SMPL
J4504-01 - 1	081418-DBRI-F-D-B	52	-0.0003	-245.1		PPB	17 Aug 2018 09:41:50	U	SMPL
J4504-02 - 1	081418-DBRI-F-D-S	96	0.0130	-245.1		PPB	17 Aug 2018 09:43:59	U	SMPL
J4504-03 - 1	081418-DBRI-F-U-B	63	0.0030	-245.1		PPB	17 Aug 2018 09:46:07	U	SMPL
J4504-04 - 1	081418-DBRI-F-U-S	100	0.0142	-245.1		PPB	17 Aug 2018 09:48:16	U	SMPL
J4504-05 - 1	081418-TIPC-F-D-B	111	0.0175	-245.1		PPB	17 Aug 2018 09:50:22	U	SMPL
CCV10 - 1	CCV10	16429	4.9497	-245.1		PPB	17 Aug 2018 09:52:29	U	SMPL
CCB10 - 1	CCB10	41	-0.0036	-245.1		PPB	17 Aug 2018 09:54:36	U	SMPL
J4504-06 - 1	081418-TIPC-F-D-S	39	-0.0042	-245.1		PPB	17 Aug 2018 09:56:44	U	SMPL
J4504-07 - 1	081418-TIPC-F-U-B	145	0.0278	-245.1		PPB	17 Aug 2018 10:01:51	U	SMPL
J4504-08 - 1	081418-TIPC-F-U-S	128	0.0227	-245.1		PPB	17 Aug 2018 10:06:20	U	SMPL
J4504-08DUP - 1	081418-TIPC-F-U-SDUP	120	0.0203	-245.1		PPB	17 Aug 2018 10:08:27	U	SMPL
J4504-08MS - 1	081418-TIPC-F-U-SMS	13345	4.0176	-245.1		PPB	17 Aug 2018 10:10:35	U	SMPL
J4504-08MSD - 1	081418-TIPC-F-U-SMSD	13478	4.0578	-245.1		PPB	17 Aug 2018 10:12:43	U	SMPL
J4504-08LX5 - 1		-116	-0.0511	-245.1		PPB	17 Aug 2018 10:14:52	U	SMPL
J4504-08A - 1		13929	4.1941	-245.1		PPB	17 Aug 2018 10:20:24	U	SMPL
CCV11 - 1	CCV11	16409	4.9437	-245.1		PPB	17 Aug 2018 10:22:32	U	SMPL
CCB11 - 1	CCB11	-94	-0.0444	-245.1		PPB	17 Aug 2018 10:24:40	U	SMPL