



TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: apatel

ANALYST: Aparna

Date: 08/16/2018

TEMP1 IN: 103 08/16/2018 15:00 TEMP OUT: 104 08/16/2018 16:10

TEMP2 IN: 104 08/16/2018 16:45 TEMP OUT: 104 08/17/2018 07:00

TEMP3 IN: 105 08/17/2018 07:45 TEMP OUT: 105 08/17/2018 09:20

TEMP4 IN: 105 08/17/2018 10:00 TEMP OUT: 105 08/17/2018 11:30

Run Number: LB97361

BalanceID: WC SC-3

OvenID: WC Oven-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L
1	LB97361BL	LB97361BL	1.4021	1.4021	100	1.4021	1.4021	1.4021	0.0000	0
2	LB97361BS	LB97361BS	1.3957	1.3957	100	1.4510	1.4510	1.4510	0.0553	553
3	J4527-01	SPDES-1	1.4008	1.4008	900	1.4020	1.4020	1.4020	0.0012	1.3

A = Sample Volume (ml)

B = Empty Dish Weight (g)

C = Final Empty Dish + Sample weight after 1.5 hr drying @105°C(g)

D = Weight (g)

Weight (g) = C - B

Result mg/L = $\frac{D}{A} \times 1000 \times 1000$

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TSS081718-2

WorkList ID : 115752

Date : 8/17/2018 8:05:19 AM

1397361

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/17/2018	Water	J4527-01	TSS	Cool 4 deg C	SKAN06	Q12	SPDES-1	08/16/2018	SM2540 D

FP-9823888

AS

07:00

Date/Time 08/17/18, 07:00

Received by: AS

Relinquished by: AS

Date/Time 08/17/18, 09:00

Received by: AS

Relinquished by: AS