



Analytical Summary Report

Analysis Method: 9040C

Analyst By : sweta

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB97362

Slope : 98.8

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			22.2	4.01	08/17/2018	09:10
2	CAL2	1	Water			22.0	7.01	08/17/2018	09:12
3	CAL3	1	Water			21.9	10.04	08/17/2018	09:14
4	ICV	1	Water			22.2	7.01	08/17/2018	09:16
5	CCV1	1	Water			21.8	2.00	08/17/2018	09:18
6	J4523-03	1	Water			23.1	9.50	08/17/2018	09:20
7	J4523-03DUP	1	Water			23.2	9.51	08/17/2018	09:22
8	CCV2	1	Water			21.9	12.00	08/17/2018	09:28

WORKLIST(Hardcopy Internal Chain)

LB 97362

WorkList Name : corrosivity081718-1

WorkList ID : 115760

Date : 8/17/2018 8:25:57 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/16/2018	Water	J4523-03	Corrosivity	Cool 4 deg C	AREC01	P33	WC-1	08/16/2018	9040C

Date/Time 08/17/18 8:30
 Received by: SP
 Relinquished by: JP

Date/Time 08/17/18 9:30
 Received by: JP
 Relinquished by: SP