



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB97364

Ambient Barometric Pressure (mmHg): 755

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	27.0	80.6	80.9	08/17/2018	09:35
2	J4516-07		1	7	>100.0	>212.0	>212.0	08/17/2018	10:05
3	J4516-07DUP		1	7	>100.0	>212.0	>212.0	08/17/2018	10:35
4	J4523-03		1	8	>100.0	>212.0	>212.0	08/17/2018	11:05

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

WorkList Name : FP81718

WorkList ID : 115767

Date : 8/17/2018 8:47:25 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/21/2018	Water	J4516-07	Flash Point	Cool 4 deg C	PSEG01	K63	ND-BASELINE-WATER	08/16/2018	1010A
08/30/2018	Water	J4523-03	Flash Point	Cool 4 deg C	AREC01	P33	WC-1	08/16/2018	1010A

Date/Time

Received by:

Relinquished by:

08/17/18 8:50 AM

KS

JP

Date/Time

Received by:

Relinquished by:

08/17/18 11:10 AM

JP

KS

LB 97364