



Analytical Summary Report

Analysis Method: 9045D

Analyst By : Sweta

Parameter: pH

Supervisor Review By : apatel

Run Number: LB97375

Slope : 98.8

BalanceID: WC SC-4

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.9	4.01	08/17/2018	13:30
2	CAL2	1	Water			22.0	7.01	08/17/2018	13:32
3	CAL3	1	Water			21.9	10.04	08/17/2018	13:34
4	ICV	1	Water			22.2	7.00	08/17/2018	13:36
5	CCV1	1	Water			21.8	2.00	08/17/2018	13:38
6	J4531-01	1	Solid	20.02	20	25.0	6.62	08/17/2018	13:40
7	J4531-01DUP	1	Solid	20.03	20	25.1	6.63	08/17/2018	13:42
8	CCV2	1	Water			22.0	12.00	08/17/2018	13:44

WORKLIST(Hardcopy Internal Chain)

LB9737S

WorkList Name : pH081718

WorkList ID : 115781

Date : 8/17/2018 1:22:57 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/16/2018	Solid	J4531-01	pH	Cool 4 deg C	TULL02	Q22	FK-PS-01-036	08/16/2018	9045D

Date/Time 08/17/18 1:25
 Received by: SP
 Relinquished by: SP

Date/Time 08/17/18 7:50
 Received by: SP
 Relinquished by: SP