



Analytical Summary Report

Analysis Method: SM5220 D

ANALYST: Indiana

Parameter: COD

SUPERVISOR REVIEW BY: apatel

Run Number: LB97379

Reagent/Standard	Lot/Log #
COD ICV-LCS std, 50ppm	WP66477
COD calibration std. 100 ppm	WP66474
COD calibration std. 10 ppm	WP66473
COD calibration std. 150 ppm	WP66476
COD calibration std. 50 ppm	WP66475
COD calibration std. 0 ppm	WP66472
COD Digestion Vials Low Level 0-150Mg/L	W2375
COD ICV-LCS std, 50ppm	WP67117
COD CCV std, 50ppm	WP67118

Temp In (C): 150	Date In: 08/17/2018	Time In: 10:00
Temp Out (C): 150	Date Out: 08/17/2018	Time Out: 12:00

Intercept: 0.1033

Slope: 1.0048

Regression: 1

Seq	Lab ID	TrueValue (mg/l)	DF	MATRIX	Reading	Anal Date	Anal Time
1	CAL1	0	1	Water	0.000	07/26/2018	12:42
2	CAL2	10	1	Water	10.000	07/26/2018	12:42
3	CAL3	50	1	Water	51.000	07/26/2018	12:42
4	CAL4	100	1	Water	100.000	07/26/2018	12:43
5	CAL5	150	1	Water	151.000	07/26/2018	12:44



Analytical Summary Report

Analysis Method: SM5220 D

ANALYST: Indiana

Parameter: COD

SUPERVISOR REVIEW BY: apatel

Run Number: LB97379

Seq	Lab ID	True Value (mg/l)	Initial Weight (g)	Final Vol (ml)	DF	MATRIX	Reading	Result	AnalDate	AnalTime
1	ICV	50			1	Water	51.000	50.654	07/26/2018	12:44
2	ICB				1	Water	0.000	-0.103	07/26/2018	12:45
3	CCV1	50			1	Water	52.000	51.649	08/17/2018	13:05
4	CCB1				1	Water	0.000	-0.103	08/17/2018	13:05
5	LB97379BL				1	Water	0.000	-0.103	08/17/2018	13:06
6	LB97379BS	50			1	Water	51.000	50.654	08/17/2018	13:06
7	J4502-03				10	Water	25.000	24.778	08/17/2018	13:07
8	J4516-06				1	Water	17.000	16.816	08/17/2018	13:07
9	J4516-06DUP				1	Water	16.000	15.821	08/17/2018	13:08
10	J4516-06MS	50			1	Water	65.000	64.587	08/17/2018	13:08
11	J4516-06MSD	50			1	Water	64.000	63.591	08/17/2018	13:09
12	CCV2	50			1	Water	52.000	51.649	08/17/2018	13:09
13	CCB2				1	Water	0.000	-0.103	08/17/2018	13:10

LB97394

WORKLIST(Hardcopy Internal Chain)

WorkList Name : COD081718 WorkList ID : 115769 Date : 8/17/2018 9:17:11 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/20/2018	Water	J4502-03	COD	Conc H2SO4 tr	PSEG01	K63	BROWN-FRAC-TANK	08/15/2018	SM5220 D
08/21/2018	Water	J4516-06	COD	Conc H2SO4 tr	PSEG01	K63	ND-BASELINE-WATER	08/16/2018	SM5220 D

Date/Time 08/17/18 9:30
Received by: JB
Relinquished by: JP

Date/Time 08/17/18 11:00
Received by: JP
Relinquished by: JB