


7471B

LB 97381
INSTRUMENT ID: CV1

Linear

 μ Abs.:

43655



A= 0.0000e+000

B= 2.2805e-004

C= 9.2377e-003

Rho= 0.9999688

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.013	0.013	16	0.000	16				
0.2	0.200	0.183	-0.017	763	0.0 %	763				
2.5	2.500	2.495	-0.005	10899	0.0 %	10899				
5.0	5.000	4.988	-0.012	21830	0.0 %	21830				
7.5	7.500	7.557	0.057	33096	0.0 %	33096				
10.0	10.000	9.965	-0.035	43655	0.0 %	43655				

LB97381 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 17 Aug 2018 15:09:57

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	16	-	0.0000	7471B	PPB	17 Aug 2018 15:12:58	S	Std
0.2 - 1	50.2	763	-	0.2000	7471B	PPB	17 Aug 2018 15:15:16	S	Std
2.5 - 1	52.5	10899	-	2.5000	7471B	PPB	17 Aug 2018 15:17:34	S	Std
5.0 - 1	55	21830	-	5.0000	7471B	PPB	17 Aug 2018 15:19:51	S	Std
7.5 - 1	57.5	33096	-	7.5000	7471B	PPB	17 Aug 2018 15:22:08	S	Std
10.0 - 1	510	43655	-	10.0000	7471B	PPB	17 Aug 2018 15:24:25	S	Std
ICV15 - 1	ICV15	17258	3.9449	-	7471B	PPB	17 Aug 2018 15:27:13	U	SMPL
ICB15 - 1	ICB15	-126	-0.0195	-	7471B	PPB	17 Aug 2018 15:29:28	U	SMPL
CCV17 - 1	CCV17	22545	5.1506	-	7471B	PPB	17 Aug 2018 15:31:44	U	SMPL
CCB17 - 1	CCB17	-133	-0.0211	-	7471B	PPB	17 Aug 2018 15:34:00	U	SMPL
CRA - 1	CRA	776	0.1862	-	7471B	PPB	17 Aug 2018 15:36:17	U	SMPL
HighStd - 1	HighStd	43734	9.9828	-	7471B	PPB	17 Aug 2018 15:38:33	U	SMPL
ChkStd - 1	ChkStd	31936	7.2922	-	7471B	PPB	17 Aug 2018 15:40:49	U	SMPL
PB112142BL - 1	PBS	-236	-0.0446	-	7471B	PPB	17 Aug 2018 15:43:08	U	SMPL
PB112142BS - 1	LCSS	16286	3.7233	-	7471B	PPB	17 Aug 2018 15:45:26	U	SMPL
J4509-02 - 1	TOD1-FP-0-4	1458	0.3417	-	7471B	PPB	17 Aug 2018 15:47:43	U	SMPL
J4511-02 - 1	RA-081418-FL-024	2647	0.6129	-	7471B	PPB	17 Aug 2018 15:50:00	U	SMPL
J4511-03 - 1	RA-081418-ESW-15NSC-025	4300	0.9899	-	7471B	PPB	17 Aug 2018 15:52:18	U	SMPL
J4511-06 - 1	RA-081418-FL-028	13792	3.1545	-	7471B	PPB	17 Aug 2018 15:54:35	U	SMPL
J4511-07 - 1	RA-081518-NSW-45-029	31007	7.0804	-	7471B	PPB	17 Aug 2018 15:56:51	U	SMPL
CCV18 - 1	CCV18	21662	4.9493	-	7471B	PPB	17 Aug 2018 15:59:08	U	SMPL
CCB18 - 1	CCB18	-176	-0.0309	-	7471B	PPB	17 Aug 2018 16:01:25	U	SMPL
J4511-08 - 1	RA-081518-FL-030	6749	1.5483	-	7471B	PPB	17 Aug 2018 16:03:42	U	SMPL
J4511-09 - 1	RA-081518-FL-031	5347	1.2286	-	7471B	PPB	17 Aug 2018 16:05:58	U	SMPL
J4511-09DUP - 1	RA-081518-FL-031DUP	4385	1.0092	-	7471B	PPB	17 Aug 2018 16:08:14	U	SMPL
J4511-10 - 1	J4511-09MS	20431	4.6685	-	7471B	PPB	17 Aug 2018 16:10:31	U	SMPL
J4511-11 - 1	J4511-09MSD	22383	5.1137	-	7471B	PPB	17 Aug 2018 16:12:47	U	SMPL
J4516-01 - 1	ND-SOIL	4338	0.9985	-	7471B	PPB	17 Aug 2018 16:15:04	U	SMPL
J4524-01 - 1	Y-1802	7681	1.7609	-	7471B	PPB	17 Aug 2018 16:17:22	U	SMPL
J4524-02 - 1	Y-1803	7175	1.6455	-	7471B	PPB	17 Aug 2018 16:19:40	U	SMPL
PB112143BL - 1	PBS	-43	-0.0006	-	7471B	PPB	17 Aug 2018 16:21:57	U	SMPL
PB112143BS - 1	LCSS	15087	3.4498	-	7471B	PPB	17 Aug 2018 16:24:13	U	SMPL
CCV19 - 1	CCV19	21613	4.9381	-	7471B	PPB	17 Aug 2018 16:26:29	U	SMPL
CCB19 - 1	CCB19	-140	-0.0227	-	7471B	PPB	17 Aug 2018 16:28:46	U	SMPL
J4521-01 - 1	RT2286	2264	0.5255	-	7471B	PPB	17 Aug 2018 16:31:03	U	SMPL
J4521-04 - 1	CHRT26100	12368	2.8298	-	7471B	PPB	17 Aug 2018 16:33:19	U	SMPL
J4521-07 - 1	72-11969	9029	2.0683	-	7471B	PPB	17 Aug 2018 16:35:36	U	SMPL
J4521-10 - 1	72-11975	2436	0.5648	-	7471B	PPB	17 Aug 2018 16:37:52	U	SMPL
J4521-13 - 1	72-12003	4180	0.9625	-	7471B	PPB	17 Aug 2018 16:40:08	U	SMPL
J4521-16 - 1	72-11990	5797	1.3312	-	7471B	PPB	17 Aug 2018 16:42:25	U	SMPL
J4521-19 - 1	72-11967	6070	1.3935	-	7471B	PPB	17 Aug 2018 16:44:43	U	SMPL
J4521-22 - 1	72-11979	6525	1.4973	-	7471B	PPB	17 Aug 2018 16:47:00	U	SMPL
J4521-25 - 1	72-11988	6811	1.5625	-	7471B	PPB	17 Aug 2018 16:49:18	U	SMPL
J4521-28 - 1	CHRT-27245	13724	3.1390	-	7471B	PPB	17 Aug 2018 16:51:34	U	SMPL
CCV20 - 1	CCV20	21970	5.0195	-	7471B	PPB	17 Aug 2018 16:53:50	U	SMPL
CCB20 - 1	CCB20	-148	-0.0245	-	7471B	PPB	17 Aug 2018 16:56:07	U	SMPL
J4523-02 - 1	WC-2	4357	1.0029	-	7471B	PPB	17 Aug 2018 16:58:24	U	SMPL
J4531-01 - 1	FK-PS-01-036	467	0.1157	-	7471B	PPB	17 Aug 2018 17:00:40	U	SMPL
J4531-01DUP - 1	FK-PS-01-036DUP	472	0.1169	-	7471B	PPB	17 Aug 2018 17:02:56	U	SMPL
J4531-01MS - 1	FK-PS-01-036MS	16480	3.7675	-	7471B	PPB	17 Aug 2018 17:05:13	U	SMPL
J4531-01MSD - 1	FK-PS-01-036MSD	17187	3.9287	-	7471B	PPB	17 Aug 2018 17:07:29	U	SMPL
J4511-09LX5 - 1		1091	0.2580	-	7471B	PPB	17 Aug 2018 17:09:47	U	SMPL
J4511-09A - 1		22576	5.1577	-	7471B	PPB	17 Aug 2018 17:12:04	U	SMPL
J4531-01LX5 - 1		-13	0.0063	-	7471B	PPB	17 Aug 2018 17:14:22	U	SMPL
J4531-01A - 1		17930	4.0982	-	7471B	PPB	17 Aug 2018 17:16:40	U	SMPL
CCV21 - 1	CCV21	22121	5.0539	-	7471B	PPB	17 Aug 2018 17:18:57	U	SMPL
CCB21 - 1	CCB21	-168	-0.0291	-	7471B	PPB	17 Aug 2018 17:21:13	U	SMPL