



Analytical Summary Report

Analysis Method: 9040C

Analyst By : Sweta

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB97386

Slope : 98.9

BalanceID: WC SC-4

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			22.0	4.01	08/17/2018	16:10
2	CAL2	1	Water			21.9	7.01	08/17/2018	16:12
3	CAL3	1	Water			22.1	10.04	08/17/2018	16:14
4	ICV	1	Water			22.0	7.00	08/17/2018	16:16
5	CCV1	1	Water			21.8	2.00	08/17/2018	16:30
6	J4274-04	1	Solid	20.02	20	25.2	8.95	08/17/2018	16:32
7	J4274-04DUP	1	Solid	20.03	20	25.3	8.96	08/17/2018	16:34
8	J4277-06	1	Solid	20.02	20	25.2	8.22	08/17/2018	16:36
9	J4277-08	1	Solid	20.02	20	25.5	8.28	08/17/2018	16:38
10	J4282-10	1	Solid	20.04	20	25.3	8.16	08/17/2018	16:40
11	J4539-01	1	Solid	20.02	20	24.6	7.61	08/17/2018	16:42
12	J4539-05	1	Solid	20.03	20	25.0	6.79	08/17/2018	16:44
13	CCV2	1	Water			22.0	12.00	08/17/2018	16:46

WORKLIST(Hardcopy Internal Chain)

LB97386

WorkList Name : corrosivity081718-2

WorkList ID : 115785

Date : 8/17/2018 2:44:28 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/17/2018	Solid	J4274-04	Corrosivity	Cool 4 deg C	PSEG05		HD-02-081718	08/17/2018	9040C
08/17/2018	Solid	J4277-06	Corrosivity	Cool 4 deg C	PSEG05		OR-03-081718-A	08/17/2018	9040C
08/17/2018	Solid	J4277-08	Corrosivity	Cool 4 deg C	PSEG05		OR-03-081718-B	08/17/2018	9040C
08/17/2018	Solid	J4282-10	Corrosivity	Cool 4 deg C	PSEG05		SU-04-081718	08/17/2018	9040C
08/17/2018	Solid	J4539-01	Corrosivity	Cool 4 deg C	PSEG01	K63	MS-RO-24231-OILY-DEBRI	08/17/2018	9040C
08/17/2018	Solid	J4539-05	Corrosivity	Cool 4 deg C	PSEG01	K63	MS-RO-27186	08/17/2018	9040C

Date/Time 08/17/18 4:00

Received by: [Signature]

Relinquished by: [Signature]

Date/Time 08/17/18 8:45:50

Received by: [Signature]

Relinquished by: [Signature]