



Extraction and Analytical Summary Report

Analysis Method: 1664A
Test: Oil and Grease
Run Number: LB97392
Analysis Date: 08/20/2018
BalanceID: WC SC-3
OvenID: EX OVEN-2

ANALYST: jignesh
REVIEWED BY: apatel
Extraction Date: 08/20/2018
Extraction IN Time: 07:50
Extraction OUT Time: 08:40

Dish #	Lab ID	Client ID	Matrix	pH	Sample Vol (ml)	Final Volume (ml)	Empty Dish Weight	Final Empty Dish Weight (g)	Silica Gel Weight (g)	Weight After Drying (g)	Final Weight After Drying (g)	Change Weight (g)	Result in ppm
1	LB97392BL	LB97392BL	WATER	2	1000	100	3.0614	3.0614	0	3.0614	3.0614	0.0000	0
2	LB97392BS	LB97392BS	WATER	2	1000	100	3.1984	3.1984	0	3.2169	3.2169	0.0185	18.5
3	LB97392BSD	LB97392BSD	WATER	2	1000	100	3.0114	3.0114	0	3.0301	3.0301	0.0187	18.7
4	J4548-01	001M-WILLETS-PT-BLVD (A	WATER	2	1000	100	3.0968	3.0968	0	3.1124	3.1124	0.0156	15.6
5	J4548-02	002M-35TH-AVE (AUG)	WATER	2	1000	100	3.0520	3.0520	0	3.0665	3.0665	0.0145	14.5



QC Batch# LB97392 Test: Oil and Grea Analysis Date: 08/20/2018

Chemicals Used:

Chemical Name	Chemical Lot #
HEXANE	W2392
pH Paper 0-14	W2373
Sodium Sulfate	EP1876
1:1 HCL	WP64228
Silica Gel	NA
Sand	NA

Standards Used:

Standard Name	Amount Used	Standard Lot #
LCSW	5.00 ML	WP64229
LCSWD	5.00 ML	WP64230
MS/MSD	NA	NA

BALANCE CALIBRATION / OVEN Dessicator Data

Analytical Balance ID # : WC SC-3

Before Analysis

0.0020 gram Balance:	<u>0.0019</u>	In OVEN TEMP :	<u>70</u>	Dessicator Time In :	<u>10:11</u>
1.0000 gram Balance:	<u>1.0005</u>	In Time:	<u>09:10</u>		
Bal Check Time:	<u>08:15</u>	Out OVEN TEMP:	<u>70</u>	Dessicator Time Out:	<u>10:45</u>
		Out Time:	<u>10:10</u>		

After Analysis

0.0020 gram Balance:	<u>0.0021</u>	In OVEN TEMP :	<u>71</u>	Dessicator Time In :	<u>11:46</u>
1.0000 gram Balance:	<u>1.0003</u>	In Time:	<u>11:15</u>		
Bal Check Time:	<u>12:42</u>	Out OVEN TEMP:	<u>71</u>	Dessicator Time Out:	<u>12:15</u>
		Out Time:	<u>11:45</u>		

2B9+392

WORKLIST(Hardcopy Internal Chain)

WorkList Name : O & G J4573

WorkList ID : 115798

Date : 8/20/2018 7:45:58 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/24/2018	Water	J4548-01 -F	Oil and Grease	Conc H2SO4 tk	TULL01	P62	001M-WILLETTS-PT-BLVD(AU)	08/16/2018	1664A
08/24/2018	Water	J4548-02 -F	Oil and Grease	Conc H2SO4 tk	TULL01	P62	002M-35TH-AVE(AUG)	08/16/2018	1664A

Date/Time 08/20/18 7:50 PM
Received by: JP
Relinquished by:

Date/Time 08/20/18 10:35 AM
Received by: JP
Relinquished by: