


7471B

LB97394
INSTRUMENT ID: CV1

Linear

 μ Abs.:

44032



A= 0.0000e+000

B= 2.2598e-004

C= -7.1385e-003

Rho= 0.9999514

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.020	-0.020	-56	0.000	-56				
0.2	0.200	0.195	-0.005	893	0.0 %	893				
2.5	2.500	2.503	0.003	11106	0.0 %	11106				
5.0	5.000	5.016	0.016	22229	0.0 %	22229				
7.5	7.500	7.563	0.063	33501	0.0 %	33501				
10.0	10.000	9.943	-0.057	44032	0.0 %	44032				

LB97394 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 20 Aug 2018 08:27:01

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-56	-	0.0000	7471B	PPB	20 Aug 2018 08:29:25	S	Std
0.2 - 1	50.2	893	-	0.2000	7471B	PPB	20 Aug 2018 08:31:43	S	Std
2.5 - 1	52.5	11106	-	2.5000	7471B	PPB	20 Aug 2018 08:34:01	S	Std
5.0 - 1	55	22229	-	5.0000	7471B	PPB	20 Aug 2018 08:36:18	S	Std
7.5 - 1	57.5	33501	-	7.5000	7471B	PPB	20 Aug 2018 08:41:09	S	Std
10.0 - 1	510	44032	-	10.0000	7471B	PPB	20 Aug 2018 08:43:27	S	Std
ICV16 - 1	ICV16	17660	3.9836	-	7471B	PPB	20 Aug 2018 08:46:18	U	SMPL
ICB16 - 1	ICB16	-124	-0.0352	-	7471B	PPB	20 Aug 2018 08:48:33	U	SMPL
CCV22 - 1	CCV22	24002	5.4168	-	7471B	PPB	20 Aug 2018 08:50:49	U	SMPL
CCB22 - 1	CCB22	-181	-0.0480	-	7471B	PPB	20 Aug 2018 08:53:05	U	SMPL
CRA - 1	CRA	869	0.1892	-	7471B	PPB	20 Aug 2018 08:55:22	U	SMPL
HighStd - 1	HighStd	44688	10.0913	-	7471B	PPB	20 Aug 2018 08:57:38	U	SMPL
ChkStd - 1	ChkStd	31030	7.0049	-	7471B	PPB	20 Aug 2018 08:59:54	U	SMPL
PB112187BL - 1	PBS	-144	-0.0397	-	7471B	PPB	20 Aug 2018 09:02:13	U	SMPL
PB112187BS - 1	LCSS	16560	3.7351	-	7471B	PPB	20 Aug 2018 09:04:31	U	SMPL
J4274-03 - 1	HD-02-081718	3430	0.7680	-	7471B	PPB	20 Aug 2018 09:06:48	U	SMPL
J4277-05 - 1	OR-03-081718-A	2854	0.6378	-	7471B	PPB	20 Aug 2018 09:09:06	U	SMPL
J4277-07 - 1	OR-03-081718-B	3300	0.7386	-	7471B	PPB	20 Aug 2018 09:11:23	U	SMPL
J4282-09 - 1	SU-04-081718	1665	0.3691	-	7471B	PPB	20 Aug 2018 09:13:41	U	SMPL
J4535-04 - 1	RA-081618-FL-036	2136	0.4755	-	7471B	PPB	20 Aug 2018 09:15:57	U	SMPL
CCV23 - 1	CCV23	23112	5.2157	-	7471B	PPB	20 Aug 2018 09:18:12	U	SMPL
CCB23 - 1	CCB23	-142	-0.0392	-	7471B	PPB	20 Aug 2018 09:20:28	U	SMPL
J4535-05 - 1	RA-081618-ESW-15-037	8040	1.8097	-	7471B	PPB	20 Aug 2018 09:22:44	U	SMPL
J4535-06 - 1	RA-081618-NSW-15-038	6533	1.4692	-	7471B	PPB	20 Aug 2018 09:25:00	U	SMPL
J4535-08 - 1	RA-081618-FL-040	2312	0.5153	-	7471B	PPB	20 Aug 2018 09:27:16	U	SMPL
J4535-10 - 1	RA-081618-FL-042	10998	2.4782	-	7471B	PPB	20 Aug 2018 09:29:32	U	SMPL
J4535-12 - 1	RA-081618-FL-044	2492	0.5560	-	7471B	PPB	20 Aug 2018 09:31:49	U	SMPL
J4537-01 - 1	G.L.-WC-UST-01	4625	1.0380	-	7471B	PPB	20 Aug 2018 09:34:05	U	SMPL
J4538-01 - 1	VSS1	1833	0.4071	-	7471B	PPB	20 Aug 2018 09:36:23	U	SMPL
J4538-02 - 1	VSS2	1961	0.4360	-	7471B	PPB	20 Aug 2018 09:38:40	U	SMPL
J4538-03 - 1	VSS3	1439	0.3180	-	7471B	PPB	20 Aug 2018 09:40:58	U	SMPL
J4538-04 - 1	VSS4	1643	0.3641	-	7471B	PPB	20 Aug 2018 09:43:15	U	SMPL
CCV24 - 1	CCV24	23108	5.2148	-	7471B	PPB	20 Aug 2018 09:45:31	U	SMPL
CCB24 - 1	CCB24	-91	-0.0277	-	7471B	PPB	20 Aug 2018 09:47:47	U	SMPL
J4539-02 - 1	MS-RO-27186	233317	52.7173	-	7471B	PPB	20 Aug 2018 09:50:04	U	SMPL
J4542-03 - 1	FK-GF-02-024-B	1343	0.2963	-	7471B	PPB	20 Aug 2018 09:54:58	U	SMPL
J4542-03DUP - 1	FK-GF-02-024-BDUP	1434	0.3169	-	7471B	PPB	20 Aug 2018 09:59:58	U	SMPL
J4542-03MS - 1	FK-GF-02-024-BMS	18006	4.0618	-	7471B	PPB	20 Aug 2018 10:02:15	U	SMPL
J4542-03MSD - 1	FK-GF-02-024-BMSD	18002	4.0609	-	7471B	PPB	20 Aug 2018 10:04:32	U	SMPL
J4546-01 - 1	BRICK-WALL-BPM	1426	0.3151	-	7471B	PPB	20 Aug 2018 10:06:49	U	SMPL
J4549-01 - 1	SD-151-20-WIPES	457	0.0961	-	7471B	PPB	20 Aug 2018 10:09:07	U	SMPL
J4542-03LX5 - 1		248	0.0489	-	7471B	PPB	20 Aug 2018 10:11:25	U	SMPL
J4542-03A - 1		18675	4.2130	-	7471B	PPB	20 Aug 2018 10:13:41	U	SMPL
CCV25 - 1	CCV25	23304	5.2590	-	7471B	PPB	20 Aug 2018 10:15:58	U	SMPL
CCB25 - 1	CCB25	-82	-0.0257	-	7471B	PPB	20 Aug 2018 10:18:14	U	SMPL
J4542-01 - 1	FK-GF-02-024-A	1425	0.3149	-	7471B	PPB	20 Aug 2018 10:20:31	U	SMPL
J4539-02DLX10 - 1		24316	5.4877	-	7471B	PPB	20 Aug 2018 10:22:47	U	SMPL
CCV26 - 1	CCV26	22828	5.1515	-	7471B	PPB	20 Aug 2018 10:25:04	U	SMPL
CCB26 - 1	CCB26	-131	-0.0367	-	7471B	PPB	20 Aug 2018 10:27:21	U	SMPL