



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB97398

Ambient Barometric Pressure (mmHg): 760

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.0	80.6	80.6	08/20/2018	09:55
2	J4539-07		1	8	>100.0	>212.0	>212.0	08/20/2018	10:20
3	J4539-07DUP		1	8	>100.0	>212.0	>212.0	08/20/2018	10:52
4	J4545-02		1	6	>100.0	>212.0	>212.0	08/20/2018	11:22

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

LB 97398

WORKLIST(Hardcopy Internal Chain)

WorkList Name : fp82018

WorkList ID : 115823

Date : 8/20/2018 9:20:36 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/22/2018	Water	J4539-07 C	Flash Point	Cool 4 deg C	PSEG01	K63	MS-BASELINE-WATER	08/17/2018	1010A
08/22/2018	Water	J4545-02 C	Flash Point	Cool 4 deg C	PSEG01	K63	RB-BASELINE-WATER	08/17/2018	1010A

Date/Time 08/20/18 9:25 AM
Received by: KS
Relinquished by: CP

Date/Time 08/20/18 11:25 AM
Received by: CP
Relinquished by: KS