


7470A

2397403
INSTRUMENT ID: CV1

Linear

 μ Abs.:

37469



A= 0.0000e+000

B= 2.6788e-004

C= 3.3613e-002

Rho= 0.9998067

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.044	0.044	39	0.000	39				
0.2	0.200	0.233	0.033	744	0.0 %	744				
2.5	2.500	2.400	-0.100	8834	0.0 %	8834				
5.0	5.000	5.055	0.055	18745	0.0 %	18745				
7.5	7.500	7.397	-0.103	27489	0.0 %	27489				
10.0	10.000	10.071	0.071	37469	0.0 %	37469				

LB97403 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 20 Aug 2018 13:15:05

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	39	-	0.0000	7470A	PPB	20 Aug 2018 13:46:54	S	Std
0.2 - 1	50.2	744	-	0.2000	7470A	PPB	20 Aug 2018 13:49:02	S	Std
2.5 - 1	52.5	8834	-	2.5000	7470A	PPB	20 Aug 2018 13:51:11	S	Std
5.0 - 1	55	18745	-	5.0000	7470A	PPB	20 Aug 2018 13:53:19	S	Std
7.5 - 1	57.5	27489	-	7.5000	7470A	PPB	20 Aug 2018 13:55:27	S	Std
10.0 - 1	510	37469	-	10.0000	7470A	PPB	20 Aug 2018 13:57:35	S	Std
ICV17 - 1	ICV17	14925	4.0317	-7470A	PPB	20 Aug 2018 14:00:16	U	SMPL	
ICB17 - 1	ICB17	-133	-0.0020	-7470A	PPB	20 Aug 2018 14:02:23	U	SMPL	
CCV27 - 1	CCV27	19051	5.1369	-7470A	PPB	20 Aug 2018 14:04:30	U	SMPL	
CCB27 - 1	CCB27	-159	-0.0090	-7470A	PPB	20 Aug 2018 14:06:37	U	SMPL	
CRA - 1	CRA	714	0.2249	-7470A	PPB	20 Aug 2018 14:08:44	U	SMPL	
HighStd - 1	HighStd	37261	10.0150	-7470A	PPB	20 Aug 2018 14:10:51	U	SMPL	
ChkStd - 1	ChkStd	25717	6.9226	-7470A	PPB	20 Aug 2018 14:12:57	U	SMPL	
PB112177BL - 1	PBW	-240	-0.0307	-7470A	PPB	20 Aug 2018 14:15:07	U	SMPL	
PB112177BS - 1	LCSW	12642	3.4201	-7470A	PPB	20 Aug 2018 14:17:16	U	SMPL	
PB112177TB - 1	PB112139TB	-111	0.0039	-7470A	PPB	20 Aug 2018 14:19:23	U	SMPL	
J4516-04 - 1	ND-SOIL	-16	0.0293	-7470A	PPB	20 Aug 2018 14:21:32	U	SMPL	
J4516-04DUP - 1	ND-SOILDUP	6	0.0352	-7470A	PPB	20 Aug 2018 14:23:40	U	SMPL	
J4516-04MS - 1	ND-SOILMS	12729	3.4434	-7470A	PPB	20 Aug 2018 14:25:49	U	SMPL	
J4516-04MSD - 1	ND-SOILMSD	14558	3.9334	-7470A	PPB	20 Aug 2018 14:27:55	U	SMPL	
CCV28 - 1	CCV28	18776	5.0633	-7470A	PPB	20 Aug 2018 14:30:03	U	SMPL	
CCB28 - 1	CCB28	-156	-0.0082	-7470A	PPB	20 Aug 2018 14:32:10	U	SMPL	
J4519-01 - 1	DRUM-18	32	0.0422	-7470A	PPB	20 Aug 2018 14:34:17	U	SMPL	
J4519-02 - 1	DRUM-14	5	0.0350	-7470A	PPB	20 Aug 2018 14:36:24	U	SMPL	
J4519-03 - 1	DRUM-17	16	0.0379	-7470A	PPB	20 Aug 2018 14:38:30	U	SMPL	
J4519-04 - 1	DRUM-15	62	0.0502	-7470A	PPB	20 Aug 2018 14:40:37	U	SMPL	
J4519-05 - 1	DRUM-13	1	0.0339	-7470A	PPB	20 Aug 2018 14:42:44	U	SMPL	
J4519-06 - 1	DRUM-16	17	0.0382	-7470A	PPB	20 Aug 2018 14:44:52	U	SMPL	
J4521-03 - 1	RT2286	33	0.0425	-7470A	PPB	20 Aug 2018 14:47:00	U	SMPL	
J4521-06 - 1	CHRT26100	30	0.0416	-7470A	PPB	20 Aug 2018 14:49:08	U	SMPL	
J4521-09 - 1	72-11969	35	0.0430	-7470A	PPB	20 Aug 2018 14:51:16	U	SMPL	
J4521-12 - 1	72-11975	40	0.0443	-7470A	PPB	20 Aug 2018 14:53:23	U	SMPL	
CCV29 - 1	CCV29	18740	5.0536	-7470A	PPB	20 Aug 2018 14:55:30	U	SMPL	
CCB29 - 1	CCB29	-127	-0.0004	-7470A	PPB	20 Aug 2018 14:57:37	U	SMPL	
J4521-15 - 1	72-12003	18	0.0384	-7470A	PPB	20 Aug 2018 14:59:45	U	SMPL	
J4521-18 - 1	72-11990	20	0.0390	-7470A	PPB	20 Aug 2018 15:01:52	U	SMPL	
J4521-21 - 1	72-11967	6	0.0352	-7470A	PPB	20 Aug 2018 15:03:59	U	SMPL	
J4521-24 - 1	72-11979	10	0.0363	-7470A	PPB	20 Aug 2018 15:06:06	U	SMPL	
J4521-27 - 1	72-11988	22	0.0395	-7470A	PPB	20 Aug 2018 15:08:13	U	SMPL	
J4521-30 - 1	CHRT-27245	25	0.0403	-7470A	PPB	20 Aug 2018 15:10:21	U	SMPL	
PB112178BL - 1	PBW	21	0.0392	-7470A	PPB	20 Aug 2018 15:12:29	U	SMPL	
PB112178BS - 1	LCSW	13653	3.6909	-7470A	PPB	20 Aug 2018 15:14:38	U	SMPL	
PB112178TB - 1	PB112141TB	-96	0.0079	-7470A	PPB	20 Aug 2018 15:16:46	U	SMPL	
J4516-07 - 1	ND-BASELINE-WATER	-5	0.0323	-7470A	PPB	20 Aug 2018 15:18:53	U	SMPL	
CCV30 - 1	CCV30	18958	5.1120	-7470A	PPB	20 Aug 2018 15:21:00	U	SMPL	
CCB30 - 1	CCB30	-158	-0.0087	-7470A	PPB	20 Aug 2018 15:23:07	U	SMPL	
J4516-07DUP - 1	ND-BASELINE-WATERDUP	-9	0.0312	-7470A	PPB	20 Aug 2018 15:25:15	U	SMPL	
J4516-07MS - 1	ND-BASELINE-WATERMS	12023	3.2543	-7470A	PPB	20 Aug 2018 15:27:23	U	SMPL	
J4516-07MSD - 1	ND-BASELINE-WATERMSD	12956	3.5042	-7470A	PPB	20 Aug 2018 15:29:30	U	SMPL	
J4523-03 - 1	WC-1	-51	0.0200	-7470A	PPB	20 Aug 2018 15:31:37	U	SMPL	
PB112179BL - 1	PBW	24	0.0400	-7470A	PPB	20 Aug 2018 15:33:45	U	SMPL	
PB112179BS - 1	LCSW	13688	3.7003	-7470A	PPB	20 Aug 2018 15:35:52	U	SMPL	
J4516-06 - 1	ND-BASELINE-WATER	-147	-0.0058	-7470A	PPB	20 Aug 2018 15:38:00	U	SMPL	
J4523-01 - 1	WC-1	10743	2.9114	-7470A	PPB	20 Aug 2018 15:40:09	U	SMPL	
J4523-01DUP - 1	WC-1DUP	11212	3.0371	-7470A	PPB	20 Aug 2018 15:42:18	U	SMPL	
J4523-01MS - 1	WC-1MS	26382	7.1008	-7470A	PPB	20 Aug 2018 15:44:26	U	SMPL	
CCV31 - 1	CCV31	18316	4.9401	-7470A	PPB	20 Aug 2018 15:46:34	U	SMPL	
CCB31 - 1	CCB31	-141	-0.0042	-7470A	PPB	20 Aug 2018 15:48:43	U	SMPL	
J4523-01MSD - 1	WC-1MSD	25313	6.8144	-7470A	PPB	20 Aug 2018 15:50:52	U	SMPL	
J4516-04LX5 - 1		-247	-0.0326	-7470A	PPB	20 Aug 2018 15:52:59	U	SMPL	
J4516-04A - 1		13611	3.6797	-7470A	PPB	20 Aug 2018 15:55:08	U	SMPL	
J4516-07LX5 - 1		-121	0.0012	-7470A	PPB	20 Aug 2018 15:57:16	U	SMPL	
J4516-07A - 1		13302	3.5969	-7470A	PPB	20 Aug 2018 15:59:23	U	SMPL	
J4523-01LX5 - 1		2028	0.5769	-7470A	PPB	20 Aug 2018 16:01:30	U	SMPL	
J4523-01A - 1		25156	6.7723	-7470A	PPB	20 Aug 2018 16:03:39	U	SMPL	
CCV32 - 1	CCV32	18681	5.0378	-7470A	PPB	20 Aug 2018 16:05:47	U	SMPL	
CCB32 - 1	CCB32	-204	-0.0210	-7470A	PPB	20 Aug 2018 16:07:58	U	SMPL	